



The Abdus Salam  
**International Centre  
for Theoretical Physics**  
[www.ictp.it](http://www.ictp.it)



School on Medical Physics for Radiation Therapy:  
**Dosimetry and Treatment Planning  
for Basic and Advanced Applications**

27 March - 7 April 2017  
*Miramare, Trieste, Italy*

# **Dosimetry Exercise**

**G. Hartmann**  
**EFOMP & German Cancer Research Center (DKFZ)**  
[g.hartmann@dkfz.de](mailto:g.hartmann@dkfz.de)

## **Calibration at a 6 MV photon beam with a linear accelerator**

Remark 1: Calibration here means:

**Determination of absorbed dose to water per 100 monitor units in a water phantom at reference conditions**

**using  
the IAEA Code of Practice TRS398**

Remark 2: Use Excel for calculation and plotting

## Objectives:

1. To learn of how to set up the measuring equipment
2. To be able to differentiate between a depth dose measurement and a calibration measurement
3. To know how a charge measurement obtained by using some monitor units has to be **manually** converted into dose in water per 100 MU under reference condition

## Introduction: General Dosimetry Formalism

- The absorbed dose to water in a water phantom for a beam of quality Q (here 6 MV photons) is obtained by the fundamental expression:

$$D_{w,Q} = M_Q N_{D,w,Q_0} k_{Q,Q_0}$$

Discussion of the meaning of the three quantities

**Q**

is the so-called **quality index** for high energy (HE) photons

The quality index  $Q$  for HE photons is defined as:

tissue–phantom ratio TPR

in water at depths of 20 and 10 g/cm<sup>2</sup>,


for a field size of 10 cm × 10 cm

and an SCD of 100 cm

$M_Q$  is the **chamber reading** (= measured charge)  
at the quality **Q** (=6 MV photon energy)

The chamber reading  $M_Q$  is obtained:

- with a water phantom
- an ionization chamber
- an electrometer



$$N_{D,w,Q_0}$$

is the **calibration factor**  
of the ionization chamber  
as given in the certificate:

Please note:

- 1) The calibration factor refers to a certain beam quality  $Q_0$  which usually is a Co-60 beam.
- 2) The calibration factor refers to reference conditions

Calibration Certificate			
Calibration laboratory for ionising radiation quantities		Calibration mark	
		000877	
		04-06	
<div style="display: flex; justify-content: space-between;"> <span>Object : Ionization chamber</span> </div>			
<div style="display: flex; justify-content: space-between;"> <span>Manufacturer : Scanditronix Wellhöfer, Germany</span> </div>			
<div style="display: flex; justify-content: space-between;"> <span>Type : CC04</span> </div>			
<div style="display: flex; justify-content: space-between;"> <span>Serial number : 6602</span> </div>			
<div style="display: flex; justify-content: space-between;"> <span>Beam quality : Co-60</span> </div>			
Absorbed dose to water calibration factor :		$N_{D,w} = 9.462 \times 10^8 \text{ Gy/C}$	
Measurement uncertainty :		$U = 2.2 \%$	
<div style="display: flex; justify-content: space-between;"> <span>Reference conditions : <math>T_0 : 20.0 \text{ }^\circ\text{C}</math></span> <span><math>p_0 : 101.325 \text{ kPa}</math></span> <span>R.H.: 50 %</span> </div>			
<p><small>The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor <math>k = 2</math>, which for a normal distribution provides a level of confidence of approximately 95%.</small></p> <p><small>The secondary standard of this laboratory is traceable to the PTB in Braunschweig (German Federal Institute of Physics and Metrology).</small></p> <p><small>Calibration reported in this certificate was carried out in accordance with the procedures described in the IAEA TRS 398 Code of Practice.</small></p>			
<div style="display: flex; justify-content: space-between;"> <span>Measuring conditions:</span> <div> <div style="display: flex; justify-content: space-between; margin-bottom: 5px;"> <span>Phantom size :</span> <span>30 cm × 30 cm × 30 cm</span> </div> <div style="display: flex; justify-content: space-between; margin-bottom: 5px;"> <span>Phantom material :</span> <span>water</span> </div> <div style="display: flex; justify-content: space-between; margin-bottom: 5px;"> <span>Source to phantom surface distance (SSD) :</span> <span>100 cm</span> </div> <div style="display: flex; justify-content: space-between; margin-bottom: 5px;"> <span>Field size at the phantom surface :</span> <span>10 cm × 10 cm</span> </div> <div style="display: flex; justify-content: space-between; margin-bottom: 5px;"> <span>Depth in phantom of the reference point of the chamber :</span> <span>5 g cm<sup>-2</sup></span> </div> <div style="display: flex; justify-content: space-between; margin-bottom: 5px;"> <span>Reference point of the IC :</span> <span>on the chamber axis at the centre of the cavity volume</span> </div> <div style="display: flex; justify-content: space-between; margin-bottom: 5px;"> <span>Chamber orientation :</span> <span>the beam axis perpendicular to the chamber axis</span> </div> <div style="display: flex; justify-content: space-between; margin-bottom: 5px;"> <span>If the chamber stem has a mark, the mark is oriented towards the radiation source</span> </div> <div style="display: flex; justify-content: space-between; margin-bottom: 5px;"> <span>Waterproof sleeve (PMMA) :</span> <span>NO</span> </div> <div style="display: flex; justify-content: space-between; margin-bottom: 5px;"> <span>Sleeve Serial Number:</span> <span>-</span> </div> <div style="display: flex; justify-content: space-between; margin-bottom: 5px;"> <span>Polarizing potential of collecting (central) electrode :</span> <span>300 V</span> </div> <div style="display: flex; justify-content: space-between; margin-bottom: 5px;"> <span>Dose rate :</span> <span>1.0 Gy min<sup>-1</sup></span> </div> <div style="text-align: center; margin-top: 5px;"> Recombination correction has not been applied </div> </div> </div>			
Date of calibration		Calibration performed by	
28.04.2006			
<div style="display: flex; align-items: center; justify-content: center;"> <div style="border-top: 1px solid black; width: 100%;"></div> <div style="text-align: left; font-size: small;">Dr. Igor Gomola</div> </div>		<div style="display: flex; align-items: center; justify-content: center;"> <div style="border-top: 1px solid black; width: 100%;"></div> <div style="text-align: left; font-size: small;">RNDr. Jozef Zeman</div> </div>	

$k_{Q,Q_0}$


is the so-called **beam quality factor**  
(**beam quality correction factor**)

Because the beam quality used at calibration ( $Q_0$ : Co-60) is **not the same** as that at the measurement ( $Q$ : 6 MV photons), this correction factor is required.

The beam quality factor is obtained from a table which is supplied with the dosimetry protocol (TRS 398).



# We use virtual equipment: 1) Simulation Program



# Virtual Equipment

further equipment:

- thermometer
- barometer

# Main preparations to be performed:

## 1. Prepare the virtual accelerator:

- set gantry angle at zero
  - set collimator angle at zero
- } select angle and  
press start continuously
- select type of radiation and energy
  - select reference field size
  - switch on the laser lines which mark the isocenter of the machine  
(use menu Options, left upper corner)

## **Main preparations to be performed:**

### **2. Prepare water phantom:**

- needs water filling
- needs adjustment of water surface to laser lines
- measure temperature and air pressure  
(see Environment, utmost right)

## **Main preparations to be performed:**

### **3. Prepare chamber:**

- adjust reference point of chamber to central ray
- position the chamber correctly to zero depth
- set correct voltage and polarity

## Some more details on the ionization chamber type to be used for the exercise:

### **PTW Farmer Type 30013**

Calibration factor:	$N = 5.233 \text{ Gy/C}$
Radius of sensitive volume:	$r = 3.1 \text{ mm}$
Voltage to be applied:	400 V
Polarity:	as used with calibration measurement

## Main steps of the beam calibration:

1. Determine the quality index  $Q$ 
  - **determine a PDD and use the depth dose method**
2. Determine the quality correction factor
  - use interpolation between table values
3. Determine charge under reference conditions at 100 monitor units (MU)
  - measure charge
  - apply correction factors
4. Finally obtain the output value, i.e. the absorbed dose in water per 100 MU at the reference point

## Note:

In high energy beams, cylindrical chambers are used for both, for

- a) depth dose measurements
- b) calibration measurements

Thus depth dose measurements and beam calibration can be performed with a single chamber type.

**However, they must be positioned in different ways:**

- a) for depth dose: **effective point** at measuring depth
- b) for calibration: **central axis** at measuring depth



# 1 Determine the quality index Q with the PDD method

Depth dose measurements with this virtual accelerator are performed in the following way:

The screenshot shows a software interface for depth dose measurements. It includes a 'Purpose of Measurement' section with radio buttons for 'Calibration' and 'Depth Dose Measurement'. Below this is the 'Depth Dose Measurement Parameters' section with input fields for 'Start depth', 'Stop depth', and 'Step increment', each followed by a unit of 'cm'. At the bottom is a 'Monitor Preselection' section with a dropdown menu currently showing 'MU' and a 'Measure' button. Four yellow callout boxes with arrows point to specific parts of the form: the first points to the 'Start depth' field, the second points to the 'Stop depth' field, the third points to the 'MU' dropdown, and the fourth points to the 'Measure' button.

**Purpose of Measurement:**

☐ Calibration ☒ Depth Dose Measurement

**Depth Dose Measurement Parameters**

Start depth:  cm

Stop depth:  cm

Step increment:  cm

**Monitor Preselection:**

**Measure**

Start depth must be greater than 0.5


Stop depth must be greater than start depth

MU required for each single depth (no continuous measurement)


Results can be copied and paste into an EXCEL file

# Example of a depth dose measurement at central ray

Water Temperature: 20.60 °C  
Air Pressure: 98.18 kPa



# 1 Determine the quality index Q with the PDD method



Example:

$$M_{10} = 7.238 \text{ nC}$$

$$M_{20} = 4.189 \text{ nC}$$

$$\Rightarrow PDD_{20,10} = \frac{M_{20}}{M_{10}} = 0.579$$

# 1 Determine the quality index $Q$ with the PDD method

Formula:  $Q \equiv TPR_{20,10} = 1.2661 \cdot PDD_{20,10} - 0.0595$

$$PDD_{20,10} = 0.579$$

$$Q \equiv TPR_{20,10} = 0.673$$

## 2 Determine the quality correction factor $k_Q$

Values of the quality correction factor  $k_Q$  are always given in tables in the dosimetry protocol as a function of Q


Therefore we needed the determination of the beam quality index Q before.

## 2 Determine the quality correction factor $k_Q$

IAEA TRS 398

CALCULATED VALUES OF  $k_Q$  FOR HIGH-ENERGY PHOTON  
BEAMS, FOR VARIOUS CYLINDRICAL IONIZATION CHAMBERS AS  
**A FUNCTION OF BEAM QUALITY  $TPR_{20,10}$**

Measured value



Quality index	0.62	0.65	0.673	0.68	0.70
PTW 30006/30013	0.997	0.994		0.990	0.988

by linear interpolation: 0.991

### 3 Determine the charge per 100 MU at reference point

- ☐ field size: 10 cm x 10 cm
- ☐ SSD: 100 cm
- ☐ phantom: water phantom
- ☐ measurement depth  
in water: 10 cm
- ☐ positioning of  
chamber: central electrode at  
measuring depth

### 3 Apply correction factors: a) Air density correction

- ❑ reference water temperature  $T_0=20^{\circ}\text{C}$
- ❑ reference air pressure (absolute!!!)  $P_0=101.325\text{ kPa}$

#### Example:

measured water temperature:  $T = 20.6^{\circ}\text{C}$

measured air pressure (absolute!!!):  $P = 98.18\text{ kPa}$

air density correction:

multiply measured M with:

$$\frac{(273.2 + T)}{(273.2 + T_0)} \frac{P_0}{P} = 1.034$$



### 3 Apply correction factors b) Saturation correction

used polarizing potential:  
saturation is 100% ???

400 V

measure charge under  
identical conditions with  
the lower voltage of 100 V

voltage	charge in nC
400.0	14.627
100.0	14.441

### 3 Apply correction factors b) Saturation correction

$$k_s = a_0 + a_1 \left( \frac{M_1}{M_2} \right) + a_2 \left( \frac{M_1}{M_2} \right)^2$$

TABLE 4.VII. QUADRATIC FIT COEFFICIENTS, FOR THE CALCULATION OF  $k_s$  BY THE “TWO-VOLTAGE” TECHNIQUE IN PULSED AND PULSED-SCANNED RADIATION, AS A FUNCTION OF THE VOLTAGE RATIO  $V_1/V_2$  [76]

$V_1/V_2$	Pulsed			Pulsed scanned		
	$a_0$	$a_1$	$a_2$	$a_0$	$a_1$	$a_2$
2.0	2.337	-3.636	2.299	4.711	-8.242	4.533
2.5	1.474	-1.587	1.114	2.719	-3.977	2.261
3.0	1.198	-0.875	0.677	2.001	-2.402	1.404
3.5	1.080	-0.542	0.463	1.665	-1.647	0.984
4.0	1.022	-0.363	0.341	1.468	-1.200	0.734
5.0	0.975	-0.188	0.214	1.279	-0.750	0.474

### 3 Apply correction factors c) Polarization correction

☐ reference polarity ?????

used polarizing potential: +400 V  
polarity effect ???

### 3 Apply correction factors c) Polarization correction

☐ reference polarity ?????

used polarizing potential: +400 V  
polarity effect ???

The polarity effect for photon beams usually is very small.

In such a case where no information on the polarity used at calibration is given, it is better **not** to perform any correction. It may be a wrong correction!

**Measurement uncertainty :**

**$U = 2.2 \%$**

**Reference conditions :**

**$T_0 : 20.0 \text{ }^{\circ}\text{C}$**

**$p_0: 101.325 \text{ kPa}$**

**R.H.: 50 %**

*The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , which for a normal distribution provides a level of confidence of approximately 95%.*

*The secondary standard of this laboratory is traceable to the PTB in Braunschweig (German Federal Institute of Physics and Metrology).*

*Calibration reported in this certificate was carried out in accordance with the procedures described in the IAEA TRS 398 Code of Practice.*

**Measuring conditions:**

Phantom size :	30 cm × 30 cm × 30 cm
Phantom material :	water
Source to phantom surface distance (SSD) :	100 cm
Field size at the phantom surface :	10 cm × 10 cm
Depth in phantom of the reference point of the chamber :	5 g·cm <sup>-2</sup>
Reference point of the IC :	on the chamber axis at the centre of the cavity volume
Chamber orientation :	the beam axis perpendicular to the chamber axis
If the chamber stem has a mark, the mark is oriented towards the radiation source	
Waterproof sleeve (PMMA) :	NO
Sleeve Serial Number:	-
Polarizing potential of collecting (central) electrode :	300 V
Dose rate :	1.0 Gy·min <sup>-1</sup>

Recombination correction has not been applied

Date of calibration

Head of the Dosimetry Laboratory

Calibration performed by

28.04.2006

*[Handwritten signature]*

### 3 Apply correction factors: Summary of all corrections

Measured charge per 100 MU                      14.627 nC

air density correction factor                      1.034

Saturation correction factor                      1.004

$$M_Q(\text{corrected}) = 14.627 \cdot 1.034 \cdot 1.004 = 15.187$$

## 4 Get calibration factor

Calibration Certificate	
Calibration laboratory for ionising radiation quantities	Calibration mark
	000877
	04-06
Object :	
Manufacturer :	
Type :	
Serial number :	
Beam quality :	Co-60
Absorbed dose to water calibration factor :	$N_{D,w} = 5.233 \times 10^7 \text{ Gy/C}$

## Final calculation

$$D_{w,Q} = M_Q N_{D,w,Q_0} k_{Q,Q_0}$$

$$M_Q(\text{corrected}) = 15.187$$

$$N_{D,w} = 5.233 \cdot 10^7 \text{ Gy/C}$$

$$k_Q = 0.991$$

$$D_{w,Q} = 0.788 \text{ Gy/100 MU}$$