



Maria Skłodowska-Curie

**National Research  
Institute of Oncology**

# Dosimetry of photon beams

**Paweł Kukołowicz**

**pawel.kukolowicz@nio.gov.pl**

Department of Medical Physics,

The Maria Skłodowska-Curie National Research Institute of Oncology

5 W.K. Roentgena st., 02-781 Warsaw, Poland

# Physics versus simplicity

## ion chamber

- Physics – Raport 277
- Simplicity – Raport 398

$$D_{water} = M \cdot N_{cal, fact} \cdot F_{cor}$$

calibration factor

what we measure (charge)

factor which corrects for changes of energy and/or for perturbation of the radiation field caused by the presence of detector

# Physics Raport 277

$$D_{water} = M_{cor} \cdot N_{D,Air} \cdot \left( \frac{S_{col}}{\rho} \right)_{water} / \left( \frac{S_{col}}{\rho} \right)_{air} \cdot P_u$$

$$N_{D,Air}$$

calibration factor in terms of dose to Air; independent on energy

$$\left( \frac{S_{col}}{\rho} \right)_{water} / \left( \frac{S_{col}}{\rho} \right)_{air}$$

ratio of dose to water and dose to air of mass stopping powers

$$P_u = P_{wall} \cdot P_{dis} \cdot P_{cav} \cdot P_{cel}$$

perturbation correction factor; ratio for Energy Q

$$M_{cor}$$

signal we measure; corrected for the actual conditions (temperature and preassure)

$$D_{water} = M_{cor} \cdot N_{D,Air} \cdot \left( \frac{L_{\Delta}}{\rho} \right)_{water} / \left( \frac{L_{\Delta}}{\rho} \right)_{air} \cdot P_u$$

$$\left( \frac{L_{\Delta}}{\rho} \right)_{water} / \left( \frac{L_{\Delta}}{\rho} \right)_{air}$$

restricted mass stopping collision power

# Simplicity: dosimetry-cook-book

## Raport 398

$$D_{water} = M_{cor} \cdot N_{D,water} \cdot k_{Q,Q_0}$$

$N_{D,water}$  calibration factor in terms of dose to water; dependent on energy

$k_{Q,Q_0}$  factor which converts calibration factor from reference energy to user energy

$M_{cor}$  signal we measure; corrected for the actual conditions (temperature and pressure)

$N_{D,water}$

for the chamber and for the energy (type of radiation and spectrum)

We receive this factor from primary or secondary standard laboratory  
**FOR ONE ENERGY!**

Most often calibration factor is established in Co60 beam, and

if in Co60 then following notation is used  $k_{Q,Q_0} = k_Q$


# $k_{Q,Q_0}$ from Raport 398

$$k_{Q,Q_0} = \frac{\left(\frac{S_{col}}{\rho}\right)_{air,Q}^w \cdot P_{u,Q}}{\left(\frac{S_{col}}{\rho}\right)_{air,Q_0}^w \cdot P_{u,Q_0}}$$

$$P_u = P_{wall} \cdot P_{dis} \cdot P_{cav} \cdot P_{cel}$$

$$\left(N_{D,water}\right)_Q = \left(N_{D,water}\right)_{Q_0} \cdot k_{Q,Q_0}$$

# Calibration procedure



Signal is measured with reference chamber and a chamber for which calibration factor is to be measured.  
**For Reference Chamber  $N_{D,w}$  calibration factor should be known.**

# Introduction

Characteristics of the studied cylindrical ionization chambers, based on the data provided by the manufacturers:

## PTW 30001



Volume:	0.6 cm <sup>3</sup>
Response:	$2 \times 10^{-8}$ C/Gy
Leakage:	$\pm 4 \times 10^{-15}$ A
Polarizing voltage:	max. 500 V
Cable leakage:	$10^{-12}$ C/(Gy $\times$ cm)
Wall material:	PMMA(C <sub>5</sub> H <sub>8</sub> O <sub>2</sub> ) <sub>n</sub>
Wall density:	1.18 mg/cm <sup>3</sup>
Wall thickness:	0.45 mm
Area density:	53 mg/cm <sup>2</sup>
Electrode:	Aluminum; 1 mm $\varnothing$ ; 21.2 mm long
Range of temperature:	+10°C .... +40°C
Range of relative humidity:	20% ... 75%
Ion collection time:	300V:0.18ms 400V:0.14ms 500V:0.11ms

## Similar 30013

## FC65-G



### Features

- Waterproof
- Air ionization chamber
- Vented through waterproof sleeve
- Fully guarded
- Includes Build-up Cap, with individual factory calibration certificate and user's guide

### Specifications

**Cavity Volume:** 0.65cm<sup>3</sup>

**Cavity Length:** 23.1 mm

**Cavity Radius:** 3.1 mm

**Wall Material:** Graphite

**Wall Thickness:** 0.073 g/cm<sup>2</sup>

**Central Electrode Material:** Aluminum


**Waterproof:** Yes



# Results

## Calibration coefficients for cylindrical ionization chambers

Box plot



Type of cylindrical ionization chambers

Descriptive statistics


Type of chambers	FC65-G	PTW 30013	PTW 30001
Sample size: n	13	82	18
Arithmetic mean value of $N_{D,w}$ [cGy/nC]	4.795	5.348	5.283
Median value of $N_{D,w}$ [cGy/nC]	4.796	5.355	5.303
Standard deviation value of $N_{D,w}$ [cGy/nC]	0.015	0.037	0.062
Standard deviation value of $N_{D,w}$ expressed as a percentage of the arithmetic mean value of $N_{D,w}$ [%]	0.32	0.70	1,16
$Q_1$ [cGy/nC]	4.759	5.328	5.210
$Q_3$ [cGy/nC]	4.806	5.372	5.329
$N_{D,w \max}$ [cGy/nC]	4.812	5.430	5.370
$N_{D,w \min}$ [cGy/nC]	4.759	5.259	5.175
$N_{D,w \max} / N_{D,w \min}$	1.01	1.03	1.04
Outliers	4.759	5.259, 5.261	none

Type of chambers	IBA FC65-G	PTW 30013	PTW 30001
Sample size: n	13	82	18
Arithmetic mean value of $N_{D,w}$ [cGy/nC]	4.795	5.348	5.283
Median value of $N_{D,w}$ [cGy/nC]	4.796	5.355	5.303
Standard deviation value of $N_{D,w}$ [cGy/nC]	0.015	0.037	0.062
Standard deviation value of $N_{D,w}$ expressed as a percentage of the arithmetic mean value of $N_{D,w}$ [%]	0.32	0.70	1,16
$Q_1$ [cGy/nC]	4.759	5.328	5.210
$Q_3$ [cGy/nC]	4.806	5.372	5.329
$N_{D,w \max}$ [cGy/nC]	4.812	5.430	5.370
$N_{D,w \min}$ [cGy/nC]	4.759	5.259	5.175
$N_{D,w \max} / N_{D,w \min}$	1.01	1.03	1.04
Outliers	4.759	5.259, 5.261	none

# Results

## Calibration coefficients for PTW 23343 and PPC05 ionization chambers

Box plot



Type of plane parallel ionization chambers

Descriptive statistics

Type of chambers	PTW 23343	PPC05
Sample size: n	44	13
Arithmetic mean value of $N_{D,w}$ [cGy/nC]	55.03	58.35
Median value of $N_{D,w}$ [cGy/nC]	54.85	58.35
Standard deviation value of $N_{D,w}$ [cGy/nC]	1.44	2.04
Standard deviation value of $N_{D,w}$ expressed as a percentage of the arithmetic mean value of $N_{D,w}$ [%]	2.61	3.50
$Q_1$ [cGy/nC]	54.40	56.86
$Q_3$ [cGy/nC]	55.99	60.08
$N_{D,w \text{ max}}$ [cGy/nC]	57.85	62.51
$N_{D,w \text{ min}}$ [cGy/nC]	51.22	55.52
$N_{D,w \text{ max}} / N_{D,w \text{ min}}$	1.13	1.13
Outliers	51.22, 51.85, 51.22	none


## Calibration coefficients for PTW 23343 and PPC05 ionization chambers

### Plane parallel chamber

Type of chambers	PTW 23343	PPC05
Sample size: n	44	13
Arithmetic mean value of $N_{D,w}$ [cGy/nC]	55.03	58.35
Median value of $N_{D,w}$ [cGy/nC]	54.85	58.35
Standard deviation value of $N_{D,w}$ [cGy/nC]	1.44	2.04
Standard deviation value of $N_{D,w}$ expressed as a percentage of the arithmetic mean value of $N_{D,w}$ [%]	2.61	3.50
$Q_1$ [cGy/nC]	54.40	56.86
$Q_3$ [cGy/nC]	55.99	60.08
$N_{D,w \max}$ [cGy/nC]	57.85	62.51
$N_{D,w \min}$ [cGy/nC]	51.22	55.52
$N_{D,w \max} / N_{D,w \min}$	1.13	1.13
Outliers	51.22, 51.85, 51.22	none

# Spectrum of photon beam Energy

## 6 MV Varian



$$\left( N_{D,water} \right)_Q$$

$$k_{Q,Q_0} = \frac{\left( \frac{S_{col}}{\rho} \right)_{air,Q}^w \cdot P_{u,Q}}{\left( \frac{S_{col}}{\rho} \right)_{air,Q_0}^w \cdot P_{u,Q_0}}$$

Monte Carlo calculation of nine megavoltage photon beam spectra using the BEAM code

Daryoush Sheikh-Bagheri<sup>a)</sup> and D. W. O. Rogers<sup>b)</sup>

TABLE 2.1. AVERAGE RESTRICTED STOPPING POWER RATIO OF WATER TO AIR,  $s_{\text{water,air}}$  FOR DIFFERENT PHOTON SPECTRA IN THE RANGE FROM  $^{60}\text{Co}$   $\gamma$  RAYS TO 35 MV X RAYS

$$k_{Q,Q_0} \cong \frac{\left(\frac{S_{col}}{\rho}\right)_{air,Q}^w}{\left(\frac{S_{col}}{\rho}\right)_{air,Q_0}^w}$$

6MV/15 MV      1,02

Photon spectrum	$s_{\text{water,air}}$
$^{60}\text{Co}$	1.134
4 MV	1.131
6 MV	1.127
8 MV	1.121
10 MV	1.117
15 MV	1.106
20 MV	1.096
25 MV	1.093
35 MV	1.084

# Dependence of spectrum on depth and position in beam




FIG. 2. Variations in photon energy spectra as a function of depth along central axis for a 6 MV 10 cm × 10 cm field.



# Dependence of spectrum on depth and position in beam




FIG. 4. Variations in out-of-field photon energy spectra as a function of distance from the central axis for 10 cm × 10 cm field.



# A synthetic spectrum parameter

- easily measured
- with small uncertainty


and

the dependence between the ratio of (restricted) mass stopping powers for water and air can be defined


$$k_{Q,Q_0} = \frac{\left( \frac{S_{col}}{\rho} \right)_{air,Q}^w \cdot P_{u,Q}}{\left( \frac{S_{col}}{\rho} \right)_{air,Q_0}^w \cdot P_{u,Q_0}}$$

# Report 398 – beam quality index

## TPR<sub>20,10</sub>



$$TPR_{20,10} = \frac{M_{water}(20, SCD = 100cm)}{M_{water}(10, SCD = 100cm)}$$



TECHNICAL REPORTS SERIES No. **398**

**Absorbed Dose Determination in  
 External Beam Radiotherapy**  
 An International Code of Practice for Dosimetry  
 Based on Standards of Absorbed Dose to Water

Sponsored by the IAEA, WHO, PAHO and ESTRO



INTERNATIONAL ATOMIC ENERGY AGENCY, VIENNA, 2000

# Report 398 – beam quality index

## TPR20/10

TPR20,10 can also be obtained from PDD (percent depth dose)

$$TPR_{20,10} = 1.2661 \cdot PDD_{20,10} - 0.0595$$

$$PDD_{20,10} = \frac{PDD(10,10,20, SSD = 100cm)}{PDD(10,10,10, SSD = 100cm)}$$

## TPR20/10

TPR20,10 can also be obtained from PDD

$$\begin{aligned} TPR_{20,10} = & -0.7898 + 0.0329 \cdot PDD(10,10,10, SSD = 100cm) \\ & - 0.000166 \cdot (PDD(10,10,10, SSD = 100cm))^2 \end{aligned}$$

# Corrections for influence quantities

$$D_{water} = M_{cor} \cdot N_{D,water} \cdot k_{Q,Q_0}$$

for preassure and temperature

$$k_{T,P} = \frac{(273.2 + T)}{(273.2 + T_0)} \cdot \frac{P_0}{P}$$

$T_0$ ,  $P_0$  are the temperature and preassure used for callibration

for humidity

no corrections for humidity are needed if the calibration factor was reffered to a relative humidity of 50% and is used in a relative humidity between 20 and 80%

# Corrections for influence quantities

$$D_{water} = M_{cor} \cdot N_{D,water} \cdot k_{Q,Q_0}$$

## Polarity effect

The influence of polarization on the chamber reading should be accounted for.

## Correction factor

$$k_{pol} = \frac{|M_+| + |M_-|}{2M}$$

M+, M- are reading obtained for positive and negative polarity  
M reading for polarity used routinely

# Polarity effect

The signal measured with the chamber changes the value, when the polarity is reversed.

Why?

For parallel chamber (easiest to understand) – in the measuring electrode more electrons are removed from the measurement area than are retained

a small positive charge is produced

when the polarity is changed, this charge is added to the measured signal

# Corrections for influence quantities

$$D_{water} = M_{cor} \cdot N_{D,water} \cdot k_{Q,Q_0}$$

## Polarity effect

User should inform the calibration laboratory on the polarization potential and polarity used in daily measurements.

After changing polarity one should wait some time to obtain a stable reading of the chamber. For some chambers it may be even 20 min.



# Corrections for influence quantities

$$D_{water} = M_{cor} \cdot N_{D,water} \cdot k_{Q,Q_0}$$

Ion recombination correction factor  $k_s$

The two voltage method is used.

Reading is measured for two voltages  $V_1$  and  $V_2$ .  $V_1/V_2 \geq 3$

Correction factor

$$k_s = a_0 + a_1 \cdot \left( \frac{M_1}{M_2} \right) + a_2 \cdot \left( \frac{M_1}{M_2} \right)^2$$

# Corrections for influence quantities

$$D_{water} = M_{cor} \cdot N_{D,water} \cdot k_{Q,Q_0}$$

$$k_s = a_0 + a_1 \cdot \left( \frac{M1}{M2} \right) + a_2 \cdot \left( \frac{M1}{M2} \right)^2$$

Coefficients from Table 9

Report 398

Two sets of data for pulsed and pulsed-scanned dose.

Almost all beams are pulsed.

$V_1/V_2$	Pulsed		
	$a_0$	$a_1$	$a_2$
2.0	2.337	-3.636	2.299
2.5	1.474	-1.587	1.114
3.0	1.198	-0.875	0.677
3.5	1.080	-0.542	0.463
4.0	1.022	-0.363	0.341
5.0	0.975	-0.188	0.214

# Ion recombination correction factor

Not all ions produced in the chamber reach the collection electrode

Some of them recombine

Factor little  $> 1$  should be used

Flash (ultra high dose rate)  $\approx 1.02$

# $k_{Q,Q_0}$ (in fact $k_Q$ )

TABLE 14. CALCULATED VALUES OF  $k_Q$  FOR HIGH ENERGY PHOTON BEAMS FOR VARIOUS CYLINDRICAL IONIZATION CHAMBERS AS A FUNCTION OF BEAM QUALITY  $TPR_{20,10}$  (adapted from Andreo [20])

Ionization chamber type <sup>a</sup>	Beam quality $TPR_{20,10}$														
	0.50	0.53	0.56	0.59	0.62	0.65	0.68	0.70	0.72	0.74	0.76	0.78	0.80	0.82	0.84
NE 2577	1.005	1.004	1.002	1.000	0.998	0.995	0.993	0.991	0.989	0.986	0.982	0.975	0.969	0.961	0.949
NE 2505 Farmer	1.001	1.001	1.000	0.999	0.997	0.994	0.991	0.988	0.984	0.980	0.975	0.967	0.959	0.950	0.937
NE 2505/A Farmer	1.005	1.003	1.001	0.997	0.995	0.990	0.985	0.982	0.978	0.974	0.969	0.962	0.955	0.947	0.936
NE 2505/3, 3A Farmer	1.005	1.004	1.002	1.000	0.998	0.995	0.993	0.991	0.989	0.986	0.982	0.975	0.969	0.961	0.949
NE 2505/3, 3B Farmer	1.006	1.004	1.001	0.999	0.997	0.991	0.987	0.984	0.980	0.976	0.971	0.964	0.957	0.950	0.938
NE 2571 Farmer	1.005	1.004	1.002	1.000	0.998	0.995	0.993	0.991	0.989	0.986	0.982	0.975	0.969	0.961	0.949
NE 2581 Farmer	1.005	1.003	1.001	0.998	0.995	0.991	0.986	0.983	0.980	0.975	0.970	0.963	0.956	0.949	0.937
PTW 23323 micro	1.003	1.003	1.000	0.999	0.997	0.993	0.990	0.987	0.984	0.980	0.975	0.967	0.960	0.953	0.941
PTW 23331 rigid	1.004	1.003	1.000	0.999	0.997	0.993	0.990	0.987	0.985	0.982	0.978	0.971	0.964	0.956	0.945
PTW 23332 rigid	1.004	1.003	1.001	0.999	0.997	0.994	0.990	0.988	0.984	0.980	0.976	0.968	0.961	0.954	0.943
PTW 23333	1.004	1.003	1.001	0.999	0.997	0.994	0.990	0.988	0.985	0.981	0.976	0.969	0.963	0.955	0.943
PTW 30001/30010 Farmer	1.004	1.003	1.001	0.999	0.997	0.994	0.990	0.988	0.985	0.981	0.976	0.969	0.962	0.955	0.943
PTW 30002/30011 Farmer	1.006	1.004	1.001	0.999	0.997	0.994	0.992	0.990	0.987	0.984	0.980	0.973	0.967	0.959	0.948
PTW 30004/30012 Farmer	1.006	1.005	1.002	1.000	0.999	0.996	0.994	0.992	0.989	0.986	0.982	0.976	0.969	0.962	0.950
PTW 30006/30013 Farmer	1.002	1.002	1.000	0.999	0.997	0.994	0.990	0.988	0.984	0.980	0.975	0.968	0.960	0.952	0.940
PTW 31002 flexible	1.003	1.002	1.000	0.999	0.997	0.994	0.990	0.988	0.984	0.980	0.975	0.968	0.960	0.952	0.940
PTW 31003 flexible	1.003	1.002	1.000	0.999	0.997	0.994	0.990	0.988	0.984	0.980	0.975	0.968	0.960	0.952	0.940

interpolation

# Interpolacja

NE 2505 A Farmer		
0.70	0.74	0.73
0.982	0.974	0.9...

# Practice – measurements of absorbed dose conventional accelerators

Water is recommended as the reference medium for measurements of absorbed dose and quality in photon beams.

The phantom should extend to at least 5 cm beyond all sides of the field size.

Cylindrical chambers are recommended.

It is not recommended to use non-waterproof chambers.

# Practice – measurements of absorbed dose conventional accelerators

TABLE 13. REFERENCE CONDITIONS FOR THE DETERMINATION OF ABSORBED DOSE TO WATER IN HIGH ENERGY PHOTON BEAMS

Influence quantity	Reference value or reference characteristics
Phantom material	Water
Chamber type	Cylindrical
Measurement depth $z_{ref}$	For $TPR_{20,10} < 0.7$ , $10 \text{ g/cm}^2$ (or $5 \text{ g/cm}^2$ ) <sup>a</sup> For $TPR_{20,10} \geq 0.7$ , $10 \text{ g/cm}^2$
Reference point of the chamber	On the central axis at the centre of the cavity volume
Position of the reference point of the chamber	At the measurement depth $z_{ref}$
SSD/SCD	$100 \text{ cm}^b$
Field size	$10 \text{ cm} \times 10 \text{ cm}^c$

# Why reference depth is 10 cm?



# Practice – measurements of percent depth dose conventional accelerators

Water is recommended as the reference medium for measurements.


The phantom should extend to at least 5 cm beyond all for sides of the field size.

Plane-parallel chambers are recommended. If cylindrical ionization chamber is used the effective measurement point should be taken into account.

Then complete PDD should be shifted towards the surface a distance equal to 0.6 internal cavity radius of the chamber. Usually we shift the chamber!

The chamber should be moved from the bottom to the surface!

# Wedge beam output factors measurements




The detector dimension in the wedge direction should be as small as possible.

Position of the chamber with respect to central axis should be checked carefully.

# Cross calibration

It is allowed to calibrate a chamber with a calibrated chamber ref in the secondary standard laboratory. Calibrated in quality  $Q_0$ .

The chambers are compared by alternately placing the chambers in a water phantom with their reference points at reference depth (10 cm).

$$N_{D,w,Q_0}^{\text{field}} = \frac{M_{\text{ref}}}{M_{\text{field}}} N_{D,w,Q_0}^{\text{ref}}$$


# Cross calibration in $Q_0$

Cross calibrated chamber factor is for  $Q_0$  radiation quality!

To have calibration factor for Co60 we have to apply  $k_{Q,Q_0}$  factor.

From  $Q$  to  $Q_0$  we use  $k_{Q,Q_0}$

From  $Q_0$  to  $Q$  we use  $1/k_{Q,Q_0}$

$$\left(N_{D,water}\right)_Q = \left(N_{D,water}\right)_{Q_0} \cdot k_{Q,Q_0}$$

TABLE 15. ESTIMATED RELATIVE STANDARD UNCERTAINTY<sup>a</sup> OF  $D_{w,Q}$  AT THE REFERENCE DEPTH IN WATER AND FOR A HIGH ENERGY PHOTON BEAM, BASED ON A CHAMBER CALIBRATION IN  $^{60}\text{Co}$  GAMMA RADIATION

Physical quantity or procedure	Relative standard uncertainty (%)
<i>Step 1: Standards laboratory<sup>b</sup></i>	
$N_{D,w}$ calibration of secondary standard at PSDL	0.5
Long term stability of secondary standard	0.1
$N_{D,w}$ calibration of the user dosimeter at the standard laboratory	0.4
<i>Combined uncertainty of step 1</i>	<i>0.6</i>
<i>Step 2: User high energy photon beam</i>	
Long term stability of user dosimeter	0.3
Establishment of reference conditions	0.4
Dosimeter reading $M_Q$ relative to beam monitor	0.6
Correction for influence quantities $k_i$	0.4
Beam quality correction $k_Q$ (calculated values)	1.0 <sup>c</sup>
<i>Combined uncertainty of step 2</i>	<i>1.4</i>
<b>Combined standard uncertainty of <math>D_{w,Q}</math> (steps 1 + 2)</b>	<b>1.5</b>

# Solid phantoms

In spite of their increasing popularity, the use of plastic phantoms is strongly discouraged for reference measurements (except for low energy X rays), as in general they are responsible for the largest discrepancies in the determination of absorbed dose for most beam types. This is mainly due to density variations between different batches and to the approximate nature of the procedures for scaling depths and absorbed dose (or fluence) from plastic to water. The density of the plastic should be measured for the batch of plastic in use rather than using a nominal value for the plastic type as supplied by the manufacturer, since density differences of up to 4% have been reported (see, for example, Ref. [65]). The commissioning of plastic phantoms in slab form should include a determination of the mean thickness and density of each slab, as well as the variation in thickness over a single slab and an investigation by radiograph for bubbles or voids in the plastic.

# Solid phantoms

TABLE 6. ELEMENTAL COMPOSITION (FRACTION BY WEIGHT), NOMINAL DENSITY AND MEAN ATOMIC NUMBER OF COMMON PHANTOM MATERIALS USED AS WATER SUBSTITUTES (*for comparison, liquid water is also included*)

	Liquid water <sup>a</sup>	Solid water WT1 <sup>a</sup>	Solid water RMI-457	Plastic water	Virtual water	PMMA <sup>a,b</sup>	Polystyrene <sup>a</sup>	Tissue equivalent plastic A-150 <sup>a</sup>
H	0.1119	0.0810	0.0809	0.0925	0.0770	0.0805	0.0774	0.1013
C		0.6720	0.6722	0.6282	0.6874	0.5998	0.9226	0.7755
N		0.0240	0.0240	0.0100	0.0227			0.0351
O	0.8881	0.1990	0.1984	0.1794	0.1886	0.3196		0.0523
F								0.0174
Cl		0.0010	0.0013	0.0096	0.0013			
Ca		0.0230	0.0232	0.0795	0.0231			0.0184
Br				0.0003				
Density (g/cm <sup>3</sup> )	1.000	1.020	1.030	1.013	1.030	1.190	1.060	1.127
$\bar{Z}^c$	6.6	5.95	5.96	6.62	5.97	5.85	5.29	5.49

Thank you for your attention!