



H4.SMR/773-12

**College on Medical Physics:
Radiation Protection and Imaging Techniques**

5 - 23 September 1994

Quality Control in Nuclear Medicine

***M. De Denaro & **S. Mattson**

*** Unità Sanitaria Locale
Trieste**

****Dept. of Radiation Physics
Malmö University Hospital
Malmö, Sweden**

Quality Control in Nuclear Medicine

Mario De Denaro

Trieste

and

* Sören Mattsson

Malmö

Quality Assurance QA

all aspects of a procedure which contribute to the quality of the results obtained

eg in nuclear medicine

- staff training
- testing of radiopharmaceuticals
- assessment of equipment performance
- reporting of clinical studies
- radiation protection is a part of all QA programs

Quality Control QC

specification, assessment, optimization and maintenance of a particular aspect

eg in nuclear medicine


- performance of the gammacamera
- " " " activity meters

"a chain is only as strong as its weakest link"

Acceptance testing - NEMA

Regular checks

- must be much simpler ($^{99}\text{Tc}^{\text{m}}$ or ^{57}Co , $\pm 10\%$ window, good phantoms
- the results must be understandable for those who make the tests.



PERFORMANCE MEASUREMENTS OF SCINTILLATION CAMERAS

RACCOMANDATIONS FOR QUALITY CONTROL AND PERFORMANCE ASSESMENT OF GAMMA CAMERAS

For the manufacturer

British Standard Institution (BSI)
National Electrical Manufacturers' Association (NEMA)
International Electrotechnical Commission (IEC)

For routine hospital use

International Atomic Energy Agency (IAEA)
Hospital Physicists' Association (HPA)
Department of Health and Social Security (DHSS)
Bureau of Radiological Health (BRH)
World Health Organisation (WHO)

NEMA NU 1-1986

NATIONAL ELECTRICAL MANUFACTURERS ASSOCIATION


PARAMETERS

1. - *INTRINSIC FLOOD FIELD UNIFORMITY*
2. - *INTRINSIC SPATIAL RESOLUTION*
3. - *SPATIAL DISTORTION*
4. - *SYSTEM SPATIAL RESOLUTION WITH AND WITHOUT SCATTER*
5. - *INTRINSIC COUNT RATE PERFORMANCE*
6. - *SYSTEM SENSIBILITY*
7. - *INTRINSIC ENERGY RESOLUTION*
8. - *DETECTOR HEAD SHIELDING LEAKAGE*
9. - *TOMOGRAPHIC SYSTEM RESOLUTION*
10. - *TOMOGRAPHIC CENTRE OF ROTATION*


INTRINSIC FLOOD-FIELD UNIFORMITY

– ACQUISITION –

POINT SOURCE GEOMETRY



FLOOD PHANTOM




ACQUISITION :

- 20% PHA window on the photopeak
- 64X64 Matrix
- Total counts 3 million counts (central pixel $> 4,000$ counts)
- Count rate $< 20,000$ cps

• INTRINSIC FLOOD-FIELD UNIFORMITY

– DATA ANALYSIS –



UFOV : Uniform Field Of View

CFOV : Central Field Of View

Smoothing by convolution with a 9 points filter function

1	2	1
2	4	2
1	2	1

INTEGRAL UNIFORMITY


(OVER WHOLE FIELD OF VIEW)

$$U_I \% = [(MAX - MIN) / (MAX + MIN)] * 100$$



DIFFERENTIAL UNIFORMITY



(OVER A 5-PIXEL DISTANCE)

$$U_D \% = [(HI - LOW) / (HI + LOW)] * 100$$




27/11/89





UNIFORMITY HISTOGRAM



UNIFORMITY %

	UFOV	CFOV
Integral	1.89	1.67
Differential	1.30	1.30
Standard dev	0.56	

X center (pix) = 31.67

Y center (pix) = 31.40

Horiz Width = 59.91

Verti Width = 44.05

UNIFORMITY MATRIX

ON LINE CORRECTION FACTORS

If the intrinsic flood-field uniformity test is over the limit of acceptability
then

PM GAIN SOFTWARE CALIBRATION

EXAMPLE

APM GAIN VARIATION SINCE LAST CALIBRATION IN % :

PM 1 =	9.3	PM 2 =	5.5	PM 3 =	3.2	PM 4 =	4.6	PM 5 =	3.2
PM 6 =	-0.5	PM 7 =	-1.3	PM 8 =	-1.7	PM 9 =	-2.0	PM 10 =	2.8
PM 11 =	-2.3	PM 12 =	-1.8	PM 13 =	-3.9	PM 14 =	2.5	PM 15 =	-1.4
PM 16 =	0.0	PM 17 =	-1.4	PM 18 =	-0.9	PM 19 =	-1.9	PM 20 =	-0.5
PM 21 =	-1.4	PM 22 =	-2.7	PM 23 =	0.5	PM 24 =	-0.4	PM 25 =	-2.3
PM 26 =	-1.8	PM 27 =	-0.9	PM 28 =	-5.5	PM 29 =	-1.4	PM 30 =	0.0
PM 31 =	0.5	PM 32 =	-0.4	PM 33 =	-0.4	PM 34 =	1.0	PM 35 =	-2.7
PM 36 =	0.5	PM 37 =	-4.4	PM 38 =	-1.8	PM 39 =	-0.9	PM 40 =	-3.1
PM 41 =	3.8	PM 42 =	-1.7	PM 43 =	-1.4	PM 44 =	-1.3	PM 45 =	1.4
PM 46 =	-1.5	PM 47 =	0.5	PM 48 =	-2.3	PM 49 =	-3.6	PM 50 =	3.5
PM 51 =	1.8	PM 52 =	-2.6	PM 53 =	-5.2	PM 54 =	-0.5	PM 55 =	-3.5
PM 56 =	-3.1	PM 57 =	-2.6	PM 58 =	-3.7	PM 59 =	7.3	PM 60 =	-0.9
PM 61 =	-3.8	PM 62 =	-1.7	PM 63 =	-0.4	PM 64 =	2.3	PM 65 =	-0.9
PM 66 =	-3.3	PM 67 =	-2.5	PM 68 =	-2.4	PM 69 =	-2.3	PM 70 =	-3.9
PM 71 =	-3.0	PM 72 =	-0.9	PM 73 =	-4.0	PM 74 =	-3.4	PM 75 =	-3.0
PM 76 =	-3.1	PM 77 =	-0.4	PM 78 =	-1.3	PM 79 =	-0.5	PM 80 =	-3.6
PM 81 =	0.5	PM 82 =	-3.3	PM 83 =	-3.4	PM 84 =	-3.0	PM 85 =	-3.7
PM 86 =	-2.1	PM 87 =	-1.7	PM 88 =	0.9	PM 89 =	-2.2	PM 90 =	3.9
PM 91 =	6.5	PM 92 =	6.3	PM 93 =	4.5	PM 94 =	7.1		

CALIBRATION DATE : 4 11 1994

POSITION # (deg) : 269





APM TABLE # : 48

CONVERGENCE CODE : TUNING HAS CONVERGED

CONTROL DATE : 7 25 1994

NEW UNIFORMITY MATRIX
ONE FOR EACH RADIOISOTOPE


CORRECTIONS CURRENTLY USED :



INTRINSIC SPATIAL RESOLUTION

— QUALITATIVE TEST —

BRH PHANTOM




Schematic section of test pattern. Hole separation (A_i) and minimal lead spacing between the holes (B_i) vary in the X direction in groups of six holes, as follows:

Group	I	II	III	IV	V	VI	VII	VIII	IX	X	XI	XII
A_i (cm)	0.4	0.45	0.5	0.55	0.6	0.65	0.7	0.75	0.8	0.85	0.9	0.95
B_i (cm)	0.15	0.2	0.25	0.3	0.35	0.4	0.45	0.5	0.55	0.6	0.65	0.7

— QUANTITATIVE TEST —

SLIT MASK





27/11/89

MED. NUC. TRIESTE

QC


FISICA

304



UNIF1

0



27 Nov 1989 at 11:31


INTRINSIC SPATIAL RESOLUTION

— QUANTITATIVE TEST —


LINE SPREAD FUNCTION


FWHM : Full Width at Half Maximum

FWTM : Full Width at Tenth Maximum




EXPERIMENTAL DATA






LINE SPREAD FUNCTIONS

150 ALONG X AXIS



LINE SPREAD FUNCTIONS



154 ALONG Y AXIS



SPATIAL DISTORTION

MEASUREMENT OF SPATIAL DISPLACEMENT

— SLIT MASK PHANTOM —



$C_1 \dots C_i = \text{FITTED PEAK CHANNELS FOR } i \text{ PROFILES}$

$$\text{i) MEAN DEVIATION} = \sum_{\text{ALL LINES}}^i |\bar{C} - C_i| / N$$


FOR BOTH X AND Y (N VALUES)

$$\text{ii) DIFFERENTIAL} = \text{MAX } |C_i - C_{i+1}|$$

i) AND ii) EXPRESSED AS % OF GFOV.


SYSTEM SPATIAL RESOLUTION WITH AND WITHOUT SCATTER

PARALLEL LINE SOURCE PHANTOM




PROFILE VALUES

Digital Image of Line Sources





Plot of counts in each pixel across the section





SYSTEM RESOLUTION



COLLIMATOR : GENERAL PURPOSE




WITH SCATTER

WITHOUT SCATTER








SYSTEM RESOLUTION WITH SCATTER



GENERAL PURPOSE




SUPER HIGH RESOLUTION





SYSTEM RESOLUTION



COLLIMATOR : SUPER HIGH RESOLUTION



DISTANCE SOURCE-COLL.
10 CM


DISTANCE SOURCE-COLL.
3 CM






INTRINSIC COUNT RATE PERFORMANCE

GEOMETRY FOR COUNT RATE CURVE




EXPERIMENTAL DATA



SYSTEM SENSIBILITY



PLANAR SOURCE PHAMTOM



$$S = \text{counts} / \text{MBq} \cdot \text{sec}$$


INTRINSIC ENERGY RESOLUTION

$$\text{RES \%} = \text{FWHM} / \text{PHOTOPEAK ENERGY}$$



DETECTOR HEAD SHIELDING LEAKAGE


SMALL CUBIC SOURCE



$$L (\%) = \frac{\text{observed count rate}}{\text{reference count rate}} * 100$$

TOMOGRAPHIC SYSTEM RESOLUTION

TOMOGRAPHIC PHANTOM



ACQUISITION PARAMETERS:

- 20% PHA window on the photopeak
- 128x128 Matrix
- x2 preprocess zoom
- Radius of rotation 150 ± 5 mm
- Transverse slice 10 ± 3 mm thick $\geq 250,000$ counts

#FANTOCCIO TOMO 16/3/93


16-03-93



TRANSVERSE (s1 = 64 th = 1)

FRONTAL (s1 = 64 th = 1)


SAGITTAL (s1 = 61 th = 1)




FANTO Tomo 16/3/93-RECON

TOMOGRAPHIC CENTRE OF ROTATION

POINT SOURCE
(5 cm from the axis of rotation)



SINOGRAM



FITTING $\implies A \sin (\vartheta + \phi) + B$

Expected COR : (matrix size + 1) / 2

offset $\implies R = ECOR - B$

$R (\vartheta)$: sinogram - fitted curve

338

5 1 2

3 1


0

6 0 255




RECONSTRUCTION FILTERS


— SPATIAL RESOLUTION —




NESSUN FILTRO




FILTRO SHEPP-LOGAN



FILTRO PARZEN



FILTRO WIENER



Quality Control of Gamma Cameras

(all results are registered in protocols)

- Daily:
- Energy window $^{99}\text{Tc}^m$
 - Background
 - Uniformity
 - Sensitivity
- } ^{57}Co planar source, 2000 counts

Every 2nd week:

- Uniformity $^{99}\text{Tc}^m$ flood source test and uniformity matrix collection
- Sensitivity same source (cps/MBq)
- Spatial resolution same source + bar quadrant phantom
- Position Linearity — " — — " — (mm/pixel)

SPECT:


- Uniformity test and uniformity matrix collection $^{99}\text{Tc}^m$ cylindrical phantom (other radionuclides if needed)
- Center of rotation $^{99}\text{Tc}^m$ point source



THE INSTITUTE OF
PHYSICAL SCIENCES
IN MEDICINE

Report No. 66

Quality Control of Gamma Cameras and Associated Computer Systems



But - quadrant phantom




Figure 4.1 A typical SPECT phantom

298

5 1 2

4 1

25

6 0 255

FANTOM012U

KFA-4

FANTOM CYLINDER

X1.50 MAY-05-94 08:34 AGE: 0 HALMO ALLN SJUKHUS / KFA-4





Figure 2. Reconstruction of a point source A) with no centre of rotation offset and B) when centre of rotation offset exists - note the image becomes a circle in the latter case (from IAEA TECDOC-317).

Activity meter ("Dose calibrator")

Before each measurement:

Background

Daily:

Measurement of a standard source (Cs-137)

Monthly:

Battery test

Zero test. Adjusted if necessary.

Background test. -"-

Measurements of three standard sources (Co-57, Cs-137 and Ra-226)

Documentation in protocols and in graphs placed close to the instrument

NaI(Tl)-detector for thyroid uptake measurements

Daily:

Energy window

Standard source of I-131

Background

NaI(Tl)-detectors for measurements on samples (“gamma counters, well detectors etc”

Before each measurement:

Energy window

Background

Standard source (Na-22)

Every 6 months:

Absolute calibration (all radionuclides measured)

Volume correction

Deadtime correction

Quality Control of Radiopharmaceuticals

Radiation considerations

- Radionuclide purity
- Radiochemical purity
- Activity

Pharmaceutical considerations

- Visual inspection
- Chemical purity
- pH


Biologic considerations

- Sterility
- Pyrogens


Radionuclide purity.



Activity meter Sample with and without Pb-shielding



Ge (NaI(Tl))



- " -



TABLE 3
Standards for BP radiopharmaceuticals

Preparation	Radioactivity Limits	Radionuclides purity	Radiochemical purity	Details of method
Iodinated [125 I] Albumin injection	85.0 to 115.0%	nmt 1% 126 I	nlt 95% HSA	ep on paper at 100V for 60 min using barbitone + acetate + octanoate + HCl buffer
Chromium [51 Cr] Edetate injection	90.0 to 110.0%	gamma spectrum same as std 51 Cr	nlt 95% Cr edetate	ep on paper at 30V/cm for 30 min using barbitone + nitrate buffer
Cyanocobalamin [57 Co] solution	85.0 to 115.0%	nmt 1% 60 Co and other nuclides	nlt 90% activity in the material purified by column and paper chromatography	cc on carboxymethylcellulose with water, extract with CCl_4 + o-cresol, then ppt with Me_2CO , dissolve in water, then pc in with butanol:acetic acid:water, elute and assay as above
Cyanocobalamin [58 Co] Solution	90.0 to 110.0%	nmt 1% 60 Co and other nuclides	nlt 90% activity in the material purified by column and paper chromatography	as above
Ferric Citrate [59 Fe] Solution	90.0 to 110.0%	gamma spectrum same as std 59 Fe	no test specified	complies with limit test for iron
Gallium [67 Ga] Citrate injection	90.0 to 110.0%	nmt 0.2% 66 Ga	no test specified	limit test for Zinc (5 ppm)
Colloidal Gold [198 Au] injection	90.0 to 110.0%	gamma spectrum same as 198 Au except that up to 10% 199 Au may be present	nlt 98% colloidal Au	ascending pc using acetone:water:HCl (70:20:10); colloidal Au stays at origin $R_f = 0$; soluble Au moves to sf and visualised by Sn(II) iodide spray
L-Selenomethionine [75 Se] injection	90.0 to 110.0%	gamma spectrum same as std 75 Se	nlt 90% in the principal spot	descending pc using butanol:water:glacial acetic acid (60:25:15); add L-methionine as carrier
Sodium chromate [51 Cr] Sterile solution	90.0 to 110.0%	gamma spectrum same as std 51 Cr	nlt 90%	ascending pc using water:ethanol:10M ammonia (125:50:25) for 2.5 hr. Cr(III) ions remain at origin

Sodium iodide [¹³¹ I] Solution	90.0 to 110.0%	gamma spectrum same as std [¹³¹ I] specific activity nlt 185Gnq/mg I ⁻	nlt 95% as iodide with R _f within 5% of std spot	ascending pc using methanol:water (75:25) and stds of 1% I ⁻ and 2% 10 ⁻³ for 2 hrs; stds visualised;
Sodium iodohip- purate [¹³¹ I] Injection	90.0 to 110.0%	gamma spectrum same as std [¹³¹ I]	nlt 96% as 2-iodohippuric acid nmt 2% as 2-iodobenzoic acid and/or iodide	tlc on silica gel GF254 using toluene:butanol:glacial acetic acid:water (80:20:4:1) for 75 min or 12 cm and stds of 2-iodohippuric acids and a mixture of 2-iodobenzoic
Sodium pertech- netate [^{99m} Tc] injection (fission)	90.0 to 110.0%	⁹⁹ Mo nmt 1 × 10 ⁻¹ % ¹³¹ I nmt 5 × 10 ⁻³ % ¹⁰³ Ru nmt 5 × 10 ⁻³ % ⁸⁹ Sr nmt 6 × 10 ⁻⁵ % ⁹⁰ Sr nmt 6 × 10 ⁻⁶ % alpha emitters nmt 1 × 10 ⁻⁷ %	nlt 95%	descending pc using methanol:water (80:20) for 2 hrs. TcO ₄ ⁻ has R _f about 0.6
Sodium pertech- netate [^{99m} Tc] Injection (non-fission)	90.0 to 110.0%	⁹⁹ Mo nmt 0.1% nmt 0.01% other radionuclides	nlt 95%	descending pc as above
Sodium phosphate [³² P] Injection	90.0 to 110.0%	beta spectrum same as standard nmt 0.0033% PO ₄ ⁻ per ³² P solution	nlt 95% nmt 37MBq	ascending pc using propanol:water: (75:25) with trichloroacetic acid and ammonia for 16 hrs and stds of H ₃ PO ₄
Technetium [^{99m} Tc] colloidal antimony sulphide injection	90.0 to 110.0%	as for Na ^{99m} TcO ₄ ⁻	nlt 95%	tlc on silica gel using 0.9% NaCl over 10-15 cm in 10 min. Plate heated at 110° for 10 min before use.
Technetium [^{99m} Tc] colloidal sulphur injection	90.0 to 110.0%	as for Na ^{99m} TcO ₄ ⁻	nlt 92%	10-15 cm, colloid remains at origin TcO ₄ ⁻ R _f = 0.6, other spots at R _f = 0.8-0.9

Note: nmt = not more than; nlt = not less than.


$^{99}\text{Tc}^m$ pharmaceuticals

pertechnetate
colloidal Tc

^{131}I —"—

iodide

Thin-layer and paper chromatography
silica gel or
glass



Column Chromatography





Fig. 5-7. Pyrogen testing employs either rabbits or amoebocytes obtained from blood of horseshoe crab. In first test, end point is rise in body temperature. In second test, end point is gelation of lysate of amoebocytes.

