




# Case histories of radiotherapy accidents & clinical consequences

Ben Mijnheer


## Accidents happen everywhere.....



#### Erroneous calibration, Exeter, UK, 1988

 Installation of a new cobalt source

 A physicist calibrated the new source



2/2/88. OP culibration of New Source Beeter Farmer 2570 with porte in water tank at depth 5.0. Water truk outs is dimensions (perspess) = 32 = 32 x ~21 cm to water sup T = 293 P = 760.3 SSD = 800 mm, 100 × 100 mm FIELD Farmer left on for 45 mins before any " Water trick filled and left to come to room tong overright. Furner readings (0.8 mine): 90,95,90.92,90.90, 90.90, 90.90 -> 90.905 (0.4 mins) 46.47, 46.40, 46.40, 46.42, 46.42 -> 46.42 Stendy state 0.4 min reading about a 44.48 Steady State Dorante 4.0D at 800 mm, 100x 100 = 2x 293.3 760 0.947 x 100 x 44.48 293 760.3 79.0 = 106.7 clyy/min 1/0.4 = 2.5 not 2 !!! Should have been 133.4 cGy/min Dore effective Times 90.905 - 2 = 44.483 2 = 44.48 = 0.0218 mins

#### What went wrong and how was it detected?

- The physicist has multiplied by the wrong factor (2.0 instead of 2.5) to achieve an equivalent exposure for one full minute. Tragically, this inaccuracy was then not recognised, possibly because the physicist was working on his own and his data may not have been checked
- Commonly afterwards only relative dose calculations are used to calculate the treatment time for an individual patient
- As a result of this calibration error, 205 patients were significantly overdosed (by about 25%) with increased morbidity and possible deaths as a consequence.
- The Institute of Physical Sciences in Medicine performed a National multicentre dosimetric comparison (external audit) and discovered this error

#### Lessons

- One clear lesson from this accident is that a calibration of a new cobalt source or linear accelerator must be checked and rechecked (and rechecked...).
- It is certainly possible to cross check a new installation by asking a colleague. It might even be sensible to repeat the calibration of a new source a month after its first use in case of contamination with other isotopes which might have unexpected patterns of decay.
- Participate in an external audit, e.g. the IAEA TLD audit system





- The centre reported a calculation error with one of its radiation machines – an orthovoltage machine (kVrange) – April 2008
- The calculation error took place during re-commissioning after the unit was moved from the General Campus to the Civic Campus – Nov 2004
- Checks of the calibration of the unit showed that measurements and calculations had been performed correctly at that moment – Nov 2007

## To make it short!

- The last step of generating output tables should have been to calculate the output for all cones from the absorbed dose to water in water under reference conditions, through the application of ratios of distance corrected in-air charge measurements and ratios of backscatter factors
- The re-commissioning covered only a 10x10 cm<sup>2</sup> field!

#### To make it short!

- Such step of generating output table d in this the been to calculate the output for all e opplitude in the absorbed dose to water in water not d to the ence conditions, through the applitude ence conditions, through the applitude ence corrected in-air charge perform which is and ratios of backscatter factors
  The re-come backscatter is a single vered only a 10x10 cm<sup>2</sup> field!

#### **Affected Patients**



- The treatment charts of all 326 patients affected were reviewed by their respective treating physicians and the patients were contacted for an immediate follow up appointment
- The error did only involve patients with basal cell and squamous cell carcinomas:
  - Patients were treated between November 2004-November 2007
  - Patients were treated at the Civic Campus
  - In some cases, patients received radiation up to 17% less than the prescribed dose
- Patients who received radiation therapy for any other type of cancer were not affected.

#### External review by experts



- The basis of this review and analysis of the events focused on the following questions:
  - Why were the incorrect output tables prepared during recommissioning?
  - Why was an independent second check not done prior to release of the output tables?
  - Why was the error not detected for 3 years?
- Root causes
  - Incorrect output tables were released for clinical use
  - Multiple significant tasks were assigned to the physicists
  - A comprehensive, independent second check was not performed

#### Staffing shortage behind radiation error: probe

Skin cancer patients at Civic given wrong dosages between 2004 and 2007

#### BY PAULINE TAM

A critical shortage of medical physicists at The Ottawa Hospital played a key role in an error that left 326 skin cancer patients with less than the recommended radiotherapy dosages to destroy their tumours, a provincial investigation into the incident has found.

low."

the mistake happened.

The staffing crunch, combined with gaps in safety and performance checks, was "a significant contributory factor" behind the improper calibration of a radiation treatment machine that went undetected for three years, the investigation concluded. The patients were treated at

the hospital's Civic campus

between November 2004 and November 2007 for basal-cell and squamous-cell carcinomas - slow-growing cancers that typically do not metastasize and are generally responsive to radiation or surgery. The hospital has since contacted all affected patients to offer them counselling and closer medical monitoring over the next five years. So far, none have shown signs of relapse, and none have filed formal complaints related to the underdosing, said Dr. Jeff Turnbull, the hospital's chief of staff.

job, another was a full-time manager and a third was on maternity leave. A fourth was on extended sick leave and a fifth had been reassigned to ra-diation safety dutes, leaving the remaining physicists over-worked. the remaining paysicits over-weak of the second s While some patients received radiation amounts that were 17-per-cent below what's normally recommended, it's not yet clear whether the lower doses necessarily resulted in less effective treatment, said Dr. Turnbull. "We have to be vigilant.

cist assigned to reinstall it made the programming error. Toutinely tested over the years, it wasn't until November 2007, when a more rigorous testing process was rolled out to can-cer hospitals across Canada, that the error was finally de-tected. We're sorry that this happened, but on the other hand, we think that the likelihood that someone is going to be seriously impacted by this is The hospital publicly dis-

tected. The higher standards forced hospitals to check multiple set-tings, or field sizes, on their machines to determine whether patients were indeed receiving their prescribed radi-ation doses. closed the error in April and asked Cancer Care Ontario to receiving their prescribed rati-and does. To constrain the prescribed rati-does retain the organization of the organization of the prescribed rati-tion setting on the organization establish that they had been calibrated preparity. The organization of the prescribed ratio of the prescription of the prescription of the province prescription of the prescription of the province prescription of the prescriptio launch an independent review. The provincial agency that oversees cancer treatment appointed a panel of Alberta specialists to look into how

Radiation:

'Excessive'

work hours

The panel found that, of the n medical physicists on staff at the time, five were unavailable for duty.

One was being trained for the

Continued from PAGE A1

See RADIATION on PAGEA2

#### **Report blames shortages**

Staffing shortages among medical physics personnel at the Ottawa Hospital contributed to a calculation error on a radiation machine affecting 326 patients over a three-year period, according to an independent review.

In a report released Thursday on the radiation treatment calculation error affecting patients with basal cell and squamous cell cancers at the Civic campus between November 2004 and November 2007, a three-member

panel found staffing shortages of medical physics personnel was "a significant contributory factor" in the miscalculation of radiation output on the relocated treatment unit.

The independent panel, led by Peter Dunscombe, the director of the medical physics department at Calgary's Tom Baker Cancer Centre, also cited "the cultural norm" at the hospital that allowed the scheduling of new treatment programs and equipment without ensuring ad-

equate medical physics staff were available.

While the dosage recalculation error happened in November 2004, the mistake wasn't discovered by hospital staff for three years.

The calculation error on a radiation machine at the Ottawa Hospital resulted in 326 skin cancer patients being underdosed during treatments, according to hospital officials.

- THE CANADIAN PRESS

Asked whether the physicist responsible for the error has

been disciplined, Dr. Turnbull Pop would only indicate that the in-dividual had since retired from resi the hospital. "I think it's important to rectos ognize this was a very, very competent individual whom 'COT

we had a great deal of faith in. we had a great deal of faith in. He was very diligent. And we've taken the approach ... BY AI Ha that, as this gentleman is retired currently, we want to just learn from this mistake and er of zani make sure that in the future, we the provide the most up-to-date thar and safe procedures for the paple tients of Ottawa." Among the changes recommended by the panel to avoid tor

T

Per

res

ma

tio

at

cl

similar mistakes in the future, the hospital has already implemented some key ones, Dr. Turnbull said. Chief among them is an in-

tha crease in the number of medical physicists on staff. Since pr the incident first came light, Ťe the hospital has hired three m more people resulting in a total of 14 physicists, Dr. Clark said.

The hospital has also introduced provisions to restrict the number of patients it treats if the number of medical physics staff falls below a certain acceptable threshold, Dr. Turnbull said. In all, 620 patients received

1,019 radiation treatments from the misprogrammed machine.

**BODY & HEALTH** For the latest news about health matters, go to our website

OTTAWACITIZENCOM /health

#### Ottawa

#### Short-staffing led to dosage error The individual who made the error on the ma

derstaffed." she said.

the General campus to the Civic campus of the Ot-tawa Cancer Centre in No-TRACEY TONG tracey.long@metronews.ca An Ottawa Hospital department was short-staffed by about half at vember 2004. It resulted in 326 skin cancer patients receiving up to 17 per cent lower-than-needed doses of radiathe time that a calculation error was made that caused hundreds of cancer patients to receive lower than required radiation doses over three years. A report by Cancer Care Ontario, ordered by the hospital, found the error occurred after an ortho-

voltage radiation treat

The report makes two recommendations — to de-velop local staffing stan-dards for medical physics support that reflect the complexity of radiation therapy, and to establish a threshold where medical physics staffing falls below that level. Hospital staff didn't discover the error for three years. There were no re-ported overdoses. At the time of the incident, staffing levels were down due to maternity leave and illnesses, said Dr. Brenda Clark, head of the that level. The hospital has acted

**Patient patients** chine, whom Turnbull called "very competent," has since retired. Turnbull doesn't expect any adverse results for pa-"We were seriously un-Although patients in a There were up to seven physicists at the time of the incident, compared to few instances were upse about the error, most were forgiving and understand-ing, said head of radiation and oncology, Dr. Laval Gritients. 14 now, she said. The report makes two "These are low-grade tumard. on both recommendations. on both recommendations, said Dr. Jeff Turnbull, the hospital's chief of staff. "Today, Ottawa Hospital medical physics staffing levels exceed currently ac-

"We're sorry this hap pened, but ... the likeli pened, but ... the likeli-hood of anyone being affected is extremely lo The patients will be seen twice a year for five years, compared to the usual cepted standards," Turn-bull said. once-a-year consultation

1.1

**Cancer error** well handled

Report says hospital reacted properly following blunder

Shall be set of the se	Dr. Jeff Turnball discusses steps taken by the hospital, at ottawasun.com/video	miscalculation. Dr. Lavel Grimard, the hos- pital's head of radiation oncol- ogy, said doctors have seen "no alarming increase" in the spread of cancer among the patients since the problem was discovered. Howevere, fit won't be for another two or three years that doctors will be able to	staffing shortages were a "sig- mificant contributory factor." At the same time, the panel believes a "cultural norm."	AMERICA pus from the Civic in 2004. A staff physicist recommis- tioned the unit after installa- tion. An annual quality review of the unit in 2007 uncovered
	forms of cancer who required reviews because of the miscal- culation, which resulted in an underdosage of up to 17% of the proscribed levels of radia- tion. There were no reported overdosages. The hospital went public about the problem in April. Doctors believe patients haven't suffered significant health effects because of the	that doctors will be able to assess the impact, if any, on the patients, Grimard said. <b>Extra visits</b> Doctors have increased patient with and the bomital	the hospital was to schedule new clinical programs and add equipment without hav- ing enough staff. The hospital has since improved its schffing, with the source acceding the accepted standard. The problem with the piece of radiation therapy equip- ment started after it was moved to the General cam-	the motion and the second seco





	review into "orthovolts" that saw 6 patients re an underd of radiatio their cance treatments
ANDRE FOR	IGET/SUN MEDIA
vic in 2004. t recommis- fter installa- puality review 07 uncovered	Dr. Peter Dunsco director of medical phy the Tom Baker Cancer I in Calgary, chaired the member expert review

combe, hysics a er Centre he three-w panel with the hospital's response. "Upon discovery of the problem, the Ottawa Hospital responded in an appropriate and conscientious manner,"

jon.willing@summedia.c

a consciencious mana Patients treated since 007 have received pr

#### From Pierre Scalliet, Brussels

Koreus.com



How could these

Lack of a QA program or mismanagement of a QA program

#### Incorrect manual parameter transfer Glasgow, Scotland

- Introduced a common data base for linacs, TPS and R/V system in 2005
- Previously all plans were calculated for
  1 Gy as prescribed dose and the number of MUs were scaled to correct the dose manually
- Now all plans were made for the correct prescribed dose

- 5th January 2006, Lisa Norris, 15 years old, started her whole CNS treatment at BOC
- The treatment plan was divided into head-fields and lower and upper spine-fields
- This is considered to be a complex treatment plan, performed about six times per year at the BOC



Output (MU/100cGy)

- Whole CNS plans still went by the "old system", where the TPS calculates MUs for 1 Gy with subsequent upscaling for dose per fx
- A "medulla planning form" was used, which is passed to treatment radiographers for final MU calculations

#### Annex 2: A blank copy of the first page of Medulla Planning FM.14.014 as used for Lisa Norris's treatment plan

BEATSON ONCOLOGY CENTRE - QA CONTROLLED DOCUMENT

MEDULLA PLANNING FORM FM.14.014 TWO SPINE FIELDS

Name:	Site:
B.O.C. No:	Unit:
Radiotherapist:	Date:
Physics:	

	Setup	Head fields isocentric; asymmetric jaws; customised shielding trays. Physics to move junction after every fractions (see over).					
	Site		ead (a)	Upper Spine (b)	Lower Spine (c)		
	Description	Right Lateral Left Lateral		Posterior	Post / Sup		
	Field Size (approx for first fractions						
	Jaw Settings	$\begin{array}{ccc} x_1 & y_1 \\ x_2 & y_2 \end{array}$	$\begin{array}{ccc} x_1 & y_1 \\ x_2 & y_2 \end{array}$				
	F.S.D.	ISOCENTRIC		100 cm	100 cm		
	Gantry Angle	90°	270°	0°	(i.e° to sup)		
	Collimators	° (i.e° Sup End Post)	° (i.e° Sup End Post)	90°	90°		
al	Floor Rotation	0°	0°	270°	270°		
	Beam Modifier	Shielding block tray code =	Shielding block tray code =	Wax compensator (a). tray code 17	Wax compensator (b). tray code 17		
	Beam Weight (%)	100% (a)	100% (a)	100% (b)	100% (c)		
	Output (MU/100cGy)						
	Dose Information	T.A.D. mid brain = 100%		spinal cord:%	spinal cord:%		
		Normalisation = %		max subcut:%			
	File Name: FMI	4014 Page Nun	ıber: 1 of: 1	Date: 11.	8.98		
	Issue Number:			Issued By:			

Output (MU/100cGy)

- HOWEVER "Planner X" let the TPS calculate the MUs for the full dose per fx – not for 1 Gy as intended
- Since the dose per fx to the head was 1.67 Gy, the MUs entered in the form were 67% too high for each of the head-fields

#### Annex 2: A blank copy of the first page of Medulla Planning FM.14.014 as used for Lisa Norris's treatment plan

BEATSON ONCOLOGY CENTRE - QA CONTROLLED DOCUMENT

MEDULLA PLANNING FORM FM.14.014 TWO SPINE FIELDS

Name:	Site:
B.O.C. No:	Unit:
Radiotherapist:	Date:
Physics:	

Setup	Head fields isocentric; asymmetric jaws; customised shielding trays. Physics to move junction after every fractions (see over).					
Site	Head (a)		Upper Spine (b)	Lower Spine (c)		
Description	Right Lateral Left Lateral		Posterior	Post / Sup		
Field Size (approx for first fractions						
Jaw Settings	$\begin{array}{ccc} x_1 & y_1 \\ x_2 & y_2 \end{array}$	$\begin{array}{ccc} x_1 & y_1 \\ x_2 & y_2 \end{array}$				
F.S.D.	ISOCE	INTRIC	100 cm	100 cm		
Gantry Angle	90°	270°	0°	(i.e° to sup)		
Collimators	° (i.e° Sup End Post)	° (i.e° Sup End Post)	90°	90°		
Floor Rotation	0°	0°	270°	270°		
Beam Modifier	Shielding block tray code =	Shielding block tray code =	Wax compensator (a). tray code 17	Wax compensator (b). tray code 17		
Beam Weight (%)	100% (a)	100% (a)	100% (b)	100% (c)		
Output (MU/100cGy)						
Dose Information		tion = %	spinal cord:%	spinal cord:%		
			max subcut:%	max subcut:%		

Date: 11.8.98

Issued By:

File Name: FM14014 | Page Number: 1 of: 1

Authorised By

Issue Number: 1

- This error was not found by the more senior planners who checked the plan
- The radiographer on the unit thus multiplied with the dose per fx a second time

ſМ

• 2.92 Gy per fx to the head

#### Annex 2: A blank copy of the first page of Medulla Planning FM.14.014 as used for Lisa Norris's treatment plan

BEATSON ONCOLOGY CENTRE - QA CONTROLLED DOCUMENT

MEDULLA PLANNING FORM FM.14.014 TWO SPINE FIELDS

Name:	Site:
B.O.C. No:	Unit:
Radiotherapist:	Date:
Physics:	

ne unit	Setup	Setup      Head fields isocentric; asymmetric jaws; customised shielding trays.        Physics to move junction after every fractions (see over).			
	Site	Head		Upper Spine	Lower Spine
		(a)		(b)	(c) .
dose	Description	Right Lateral	Left Lateral	Posterior	Post / Sup
	Field Size (approx for first fractions				
	Jaw Settings	$\begin{array}{ccc} x_1 & y_1 \\ x_2 & y_2 \end{array}$	x <sub>1</sub> y <sub>1</sub> x <sub>2</sub> y <sub>2</sub>		
	F.S.D.		A2 y2 ENTRIC	100 cm	100 cm
ead	Gantry Angle	90°	270°	0°	(i.e° to sup)
Jau	Collimators	° (i.e° Sup End Post)	° (i.e° Sup End Post)	90°	90°
	Floor Rotation	0°	0°	270°	270°
	Beam Modifier	Shielding block tray code =	Shielding block tray code =	Wax compensator (a). tray code 17	Wax compensator (b). tray code 17
	Beam Weight (%)	100% (a)	100% (a)	100% (b)	100% (c)
	Output (MU/100cGy)				
	Dose Information	T.A.D. mid brain = 100%		spinal cord:%	spinal cord:%
Output		Normalisation = %		max subcut:%	max subcut:%
U/100cGy)	File Name: FMI	14014 Page Num	ıber: 1 of: 1	Date: 11.8	3.98
	Issue Number: ]	1 Authorised By:		Issued By:	

#### Discovery of the accident

- "Planner X" calculated another plan of the same kind and made the same mistake
- This time, the error was discovered by a senior checker (1st of Feb ''06)
- The same day, the error in the calculations for Lisa Norris was also identified

#### Impact of the accident

- The total dose to Lisa Norris from the right and left lateral head fields was 55.5 Gy (19 x 2.92 Gy)
- She died nine months after the accident
  - Probably due to recurring disease



#### Lessons to learn


- The experienced planner supervised and checked the plan her-/himself)
- No instructions for putting values into the old form
- Could have been avoided by an independent check of the number of MUs
- In vivo dosimetry may have identified the erroneous dose
- Lack of staff (6-7000 patient annually) may have contributed

#### Lessons to learn

- Ensure that all staff
  - are properly trained in safety of critical procedures
  - are included in training programmes and has supervision as necessary, and that records of training are kept up-to-date
  - understand their responsibilities
- Include in the quality assurance program
  - formal procedures for verifying the risks following the introduction of new technologies and procedures
  - to perform independent MU checking of all treatment plans
- Review staffing levels and competencies

From Pierre Scalliet, Brussels


#### The incident triangle



#### Inappropriate beam calibration France 2007

- Reported in 2007 at Hôpital de Rangueil in Toulouse, France
- In April 2006, the physicist in the clinic commissioned the new BrainLAB Novalis stereotactic unit
  - This unit can operate with microMLCs (3 mm leaf-width) or conical standard collimators





## Background

- Very small fields can be defined with the microMLCs
  - High dose to a 6 mm x 6 mm field is within capability
  - The TPS requires percent depth doses, beam profiles and relative scatter factors down to this field size
  - Care must be taken when measuring small fields!
- Different measuring devices were used by the physicist
  - A measuring device not suitable for calibrating the smallest microbeams was used
  - "...an ionisation chamber of inappropriate dimensions..." according to Nuclear Safety Authority (ASN) inspectors
- The incorrect data was entered into the TPS
  - All patients treated with micro MLC were planned based on this incorrect data
  - Patients treated with conical collimator were not affected

#### Discovery of the accident

- The BrainLAB company discovered that the measurement files did not match up with those at other comparable centres, during a worldwide intercomparison study
- It should be noted that the company does not validate or hold any responsibility for local measurements or implementation

#### Impact of the accident

- Treatment based on the incorrect data went on for a year (Apr´06 – Apr´07)
- All patients treated with microMLCs were affected (145 of 172 stereotactic patients)
- The dosimetric impact was evaluated as small in most cases, with 6 patients identified for whom over 5% of the volume of healthy organs may have been affected by dose exceeding limits

#### Lessons to learn

#### • Ensure that staff

- understands the properties and limitations of the equipment they are using

#### • Include in the quality assurance programme

- an intercomparison with other hospitals, in this case an independent check of the output of a new accelerator by an independent group (using their own equipment) before the equipment is clinically used

**IAEA** raining Course

## Incorrect IMRT planning/delivery

# USA, NY – 2005 Discussed in **The New Hork Times** 2010



# The New York Times January 2010

- Several articles in the NYT early 2010
- Lot's of fuzz in the radiotherapy community
- Hearing in the US Senate
- Many meetings in the US on radiation safety


THE RADIATION BOOM Radiation Offers New Cures, and Ways to Do Harm

By WALT BOGDANICH Published: January 23, 2010

As Scott Jerome-Parks lay dying, he clung to this wish radiation overdose — which left him deaf, struggling t to swallow, burned, with his teeth falling out, with <u>ulc</u> mouth and throat, nauseated, in severe pain and final breathe — be studied and talked about publicly so tha not have to live his nightmare.



🕄 Enlarge This Image



buckets of sand from the beach where they had played as children so he could touch it, feel it and remember better days.

Mr. Jerome-Parks died several weeks later in 2007. He was 43.

A New York City hospital treating him for tongue <u>cancer</u> had failed to detect a computer error that directed a linear accelerator to blast his brain stem and neck with errant beams of radiation. Not once, but on three consecutive days.

For his last Christmas, Scott Jerome Parks rested his feet in buckets of sand his friends had sent from a childhood beach. More Photos »

#### Energy and Commerce - Subcommittee on Health held a hearing entitled "Medical Radiation: An Overview of the Issues" on Friday, February 26, 2010






http://www.youtube.com/watch?v=NcqRgVqeQSg http://www.youtube.com/watch?v=L\_IzTqhghMs

- Tuesday March 8, 2005
  - The patient begins an IMRT treatment at St Vincent's Hospital, Manhattan, NY
  - The plan had passed the QC process according to the local protocol
  - The treatment is delivered correctly
- Friday March 11, 2005
  - The physician reviews the case after 4 fractions
    - Wants a modified dose distribution (reducing dose to teeth)
- Monday March 14, 2005
  - Re-planning and re-optimization starts
  - Fractionation is changed. Existing fluences are deleted and reoptimized. New optimal fluences are saved
  - Final calculations are started, where MLC motion control points for IMRT are generated

- "Save all" is started; all new and modified data should be saved
- In this case, data to be saved included
  - actual fluence data
  - a DRR
  - the MLC control points






#### The transaction error message displayed





- Monday March 14, 2005, 11.a.m.
  - Within 12 s, another workstation, WS1, is used to open the patient plan. The planner would have seen this:



• Monday - March 14, 2005, 11.a.m.

No MLC control point data is included in the plan, neither required for dose calculation, display and approval !!!







The sagittal view should have looked like the one to the right, with MLCs

T Knöös

- Monday March 14, 2005, 1 p.m.
  - The patient is treated. The console screen would have indicated that the MLC is not used during treatment:





#### **Discovery of accident**

- Monday March 14, 2005, 11 a.m.
  - No verification plan is generated or used - should be done according to local QA program
  - The plan is subsequently prepared for treatment (treatment scheduling, image scheduling, etc
- It is also approved by a physician
- According to local QA program, a second physicist should then have reviewed the plan
  - including an overview of the irradiated area outline
  - MLC shape
  - Etc

- Tuesday/Wednesday March 15-16, 2005
  - The patient is treated without MLCs for three fractions
- Wednesday March 16, a verification plan is created and run on the treatment machine. The operator notices the absence of MLCs.
  - A second verification plan is created and run with the same result
- The patient received 13 Gy per fraction for three fractions, i.e. 39 Gy in 3 fractions

#### Lessons to learn

- Do what you should be doing according to your QA programme
  - The error could have been found through a verification of the plan (normal QA procedure at the facility) or an independent review
- Be alert when a computer crashes or freezes; check the data before you continue working
- Work with awareness at the treatment unit, and keep an eye out on unexpected behaviour of the machine




Mr. Jerome-Parks with his wife, Carmen, on the day he received his diagnosis of tongue cancer. For his treatment, he chose St. Vincent's Hospital in Manhattan, which was promoting a new linear accelerator and a treatment called Intensity Modulated Radiation Therapy, which could more precisely shape and modulate the radiation beam. Treatment started March 8, 2005




Sensing that death was near, Mr. Jerome-Parks and his wife summoned his family for a final Christmas together. Friends sent buckets of sand from the beach in Gulfport, Miss., where they had played together, so that he could sink his feet in it and remember happy times. Two month later in Febr. 2007 he died from his injuries.

#### Swiss cheese model of failure propagation



#### Radiotherapy safety layers



#### Summary and lessons to learn

- Work with awareness and alertness
  - Be aware of what you are doing
  - An irradiation can't be undone
- Procedures
  - Think through if procedures are covering everything that might go wrong
- Training and understanding
  - Have a thorough understanding of equipment and the data that is used for patient treatments
- Responsibilities
  - Make sure all responsibilities are allocated and understood
  - All members of staff are educated according to their tasks and kept up-to-date in their training

Thanks for listening

and thanks particularly to Tommy Knöös for using many of his slides in this presentation



#### To ensure patient safety we need

