Incident Reporting Systems

Ola Holmberg, PhD

Head, Radiation Protection of Patients Unit Radiation Safety and Monitoring Section NSRW International Atomic Energy Agency - IAEA Vienna, Austria



Contents

- **1. Background / Context**
- 2. Why do we need safety reporting and learning?
- **3.** What is the role of an event reporting system?
- 4. Purposes of mandatory event reporting
- **5.** Purposes of voluntary event reporting
- 6. Internal and external event reporting in practice
- 7. Incidents and near-misses in practice



1. Starting point: Radiotherapy has significant global importance

- An estimated 5.1 million courses of radiotherapy treatment were administered annually between 1997 and 2007 (up from an estimated 4.3 million in 1988)*
 - 50-60% of cancer patients could benefit from radiation therapy
 - The fraction of cancer patients treated is increasing, where RT is available





* UNSCEAR 2008 Report

2. Safety in radiotherapy is crucial



Short-staffing led to dosage error

staffed by about half at the time that a ital found the

ember 2004. It resulted in 326 skin physicists at the time the incident, compared

SKY NEWS

Teen May Die After Radiation Blunder

A teenager faces an uncertain future after a top cancel

Radiation Errors Reported in Missouri

By WALT BOGDANICH and REBECCA R. RUIZ Published: February 24, 2010

A hospital in Missouri said Wednesday that it had overradiated 76 patients, the vast majority with brain cancer, during a five-year period because powerful new radiation equipment had been set up incorrectly even with a representative of the manufacturer watching as it was done



daughter in the challenges ahead.

ly fatal overdose of s given the overdose 17 atment for a brain tumou e in Glasgow.

en blamed for the mistakes which happened a

ov sessions. under are not yet known - but doctors have w could be devastating

"I've got burns down the back of my neck and ally sore and I can't really do much he future holds for Lisa and she said:"I could

orani uamageu, i could be paralysed, I could not be here, time wil The problems with radiotherapy came after Lisa had undergone a operation and a course of chemotherapy. Professor Alan Rodger, medical director of Beatson Oncology Ce

said: "Initial meetings have taken place with the girl and her family "We will do everything in our power to support both them and thei



Report s

JONW

The New Hork Eimes

WORLD U.S. N.Y. / REGION BUSINESS TECHNOLOGY SCIENCE HEALTH SPORTS OPINION

RESEARCH FITNESS & NUTRITION MONEY & POLICY VIEWS

Go

Health

RECOMMEND

E TWITTER

B SEND TO

SINGLE PAGE

E-MAIL

SHARE

Search Health 3,000+ Topics

THE RADIATION BOOM

Radiation Offers New Cures, and Ways to Do Harm By WALT BOGDANICH

Published: January 23, 2010

As Scott Jerome-Parks lay dying, he clung to this wish: that his fatal radiation overdose - which left him deaf, struggling to see, unable to swallow, burned, with his teeth falling out, with ulcers in his mouth and throat, nauseated, in severe pain and finally unable to breathe be studied and talked about publicly so that others might not have to live his nightmare.



@ Enlarge This Image Sensing death was near, Mr. Jerome-Parks summoned his family for a final Christmas. His friends sent two buckets of sand from the beach where they had played as children so he could touch it, feel it and remember better days.

> Mr. Jerome-Parks died several weeks later in 2007. He was 43

A New York City hospital treating him for tongue cancer had failed to detect a computer error that directed a linear



Accueil » Actu » A la Une Radiothérapie. Après les accidents de Toulouse at Eninal Los rapports d'inspection bientôt en



ucléaire (ASN) va mettr à disposition du public" sur son site internet, les onclusions des nspections réalisées lans les centres de adiothérapie à la suite lo l'accident d'Epipal a le l'accident d'Epinal, a-t elle indiqué mardi dans in communiqué.

ASN mettra à l'été, sur n site www.asn.fr, ensemble des lettres de suite des inspections qu'elle a réalisées en 2007 et 2008 dans les

, comme elle le fait déjà pour les installations nuclé

ctions menées en 2007 dans les centres de radiothérapie, disponible sur une situation contrastée², souligne l'ASN : "il existe des centres dont la des traitements est bien initiée, voire bien avancée; d'autres centres, moins entre des l'hierces prognationnelles cui d'acquiende de corriors no contre des faiblesses organisationnelles qu'il conviendra de corriger er

e ces centres doit être réalisée d'ici fin juin "afin d'évaluer les premières actior

treatment error

REPRINTS WATCH THE NEW TRAILER O



cancer error: seven patients may have ared harm (ABC)





2. Safety in radiotherapy is crucial

- Over the last three decades, at least 3000 patients have been affected by radiotherapy incidents and accidents
- Radiation accidents involving medical uses have accounted for more acute radiation deaths than any other source, including Chernobyl
- These accidents do not only affect patients directly (e.g. harm and death), but might also undermine the public's confidence in the treatment
- Preventable medical errors overall also cost countries billions of dollars each year



Statute of the International Atomic Energy Agency:

- 1. Came into force on 29 July 1957
- "The Agency shall seek to accelerate and enlarge the contribution of atomic energy to peace, <u>health</u> and prosperity throughout the world."
- "To establish or adopt ... standards of <u>safety</u> for protection of health and minimization of danger to life ... and to provide for the application of these <u>standards</u>"







Protection from Risks



Bastion

City wall



Protection from Risks





Protection from Risks





France 2007 (1-year period)



« Farmer » chamber : 0,65 cm³
 « Pinpoint » chamber : 0,03 cm³

From: S. Derreumaux, IRSN, France





USA 2009 (5-year period)

Radiation Errors Reported in Missouri

By WALT BOGDANICH and REBECCA R. RUIZ Published: February 24, 2010

A hospital in Missouri said Wednesday that it had overradiated 76 patients, the vast majority with brain <u>cancer</u>, during a five-year period because powerful new radiation equipment had been set up incorrectly even with a representative of the manufacturer watching as it was done.

From: W. Bogdanich, N.Y.Times, USA





France 2004



From: S. Derreumaux, IRSN, France



USA 2009?

The New York Times

NEW YORK, WEDNESDAY, DECEMBER 29, 2010

A Pinpoint Beam Strays Invisibly, Harming Instead of Healing An Incorrect Setting Leads to Injury Problems involving machines that deliver therapeutic radiati

have led to patient injuries.

By WALT BOGDANICH and KRISTINA REBELO The initial accident report of-fered few details, except to say that an unidentified hospital had administered radiation over doministered radiation over-doses to three patients during identical medical procedures. It was not until many months later that the full import of what had happened in the hospital last year began to surface in urgent nationwide warnings, which advised doctors to be extra vigilant when using a particular device that delivers high-intensity, pin-point radiation to vulnerable parts of the body.

Marci Faber was one of the three patients. She had gone to Evanston Hospital in Illinois seeking treatment for pain emanating from a nerve deep inside her head. Today, she is in a nurs-ing home, nearly comatose, un-able to speak, eat or walk, leaving her husband to care for their

three young daughters. Two other patients were over-dosed before the hospital realized tor, had inexplicably allowed ra-diation to spill outside a heavy metal cone attachment that was diation to spill outside a heavy metal cone attachment that was supposed to channel the beam to target tiny tumors and other

MIKA GRÖNDAHL AND BILL MARSH/THE NEW YORK TIMES a specific spot in the brain. One anomalies affecting the brain or month later, the same accident happened at another hospital. The treatment Ms. Faber re

spinal cord, while minimizing damage to surrounding tissue. Because the radiation is so conthat the device, a linear accelera- ceived, stereotactic radiosurgery, centrated and intense, accuracy is especially important. Yet, ac-cording to records and inter-views, the SRS unit at Evanston lacked certain safety features, in-



Marci Faber is nearly coma tose after a treatment mistake

THE RADIATION BOOM Missing the Target

cluding those that might have ed radiation from leaking prevented radiation from leaking outside the cone. The mistakes in Evanston in-

volve linear accelerators - com volve intear accelerators — com-monly used for standard radia-tion therapy — that were re-designed by the manufacturer, Varian Medical Systems, so they could also perform SRS. As the devices became more versatile and complex, problems arose when vital electronic components could not communicate with one

another. In the last five years, SRS sys Continued on Page A12

From: W. Bogdanich, N.Y.Times, USA

INCORRECT SETUP

The beam passes through a

exceeding the cone's diameter

and irradiates healthy tissue,

mistakenly large opening.

causing injury.



CORRECT SETUP A beam passes through an adjustable opening and then through a heavy metal cone that focuses the beam on the treatment area.



Accidents and incidents still tend to "repeat themselves" – i.e. we need to be better at learning from previous events



Actions

Browse Safety Info by Process Step > Search for Incident Reports > Search for Documents & Links > View Instructions >

Featured Incident Reports

ROSIS Incident Report #20: Treatment with soft wedges in the wrong direction

During the treatment planning process, the field names of two tangential breast fields (e.g. left medial and left lateral tangential fields) were reversed...

ROSIS Incident Report #314: Incorrect manual entry of soft wedge direction Effective doses in radiology and diagnostic nuclear medicine.

Featured Documents & Links

Official report on the radiotherapy accident in Epinal (France)

Radiation dose and image quality for paediatric interventional cardiology.

Consectuer pelentesque erum

Effective doses in radiology and diagnostic nuclear medicine.

Safety in radiotherapy requires many safety-layers

Implementing lessons learned from reported events is only one of these layers

Initiating events



Event reporting

What is the role of an event reporting system?

An event reporting system can play an important role in ...

- identifying system design flaws and safety critical steps in the radiotherapy pathway
- highlighting critical problems and patterns of causes of these problems
- spreading knowledge on new risks or involving new technology
- promoting safety culture and safety awareness through involvement of and feedback to staff and managers

To fulfil this role, the event reporting needs to be a link in a longer chain:

 Incident Identification => Reporting => Investigation => Analysis => Management => Learning



Event reporting

What makes incident reports meaningful?

"the narrative"

Charles Billings (the designer of the Aviation Safety Reporting System in the USA)



Event reporting

Mandatory event reporting systems:

• Reporting of certain events is required (e.g. reporting to regulatory authorities on events above certain magnitude)

Voluntary event reporting systems:

• Reporting is encouraged (e.g. reporting to professional organization or international organization, voluntarily)

Internal event reporting systems:



• Reporting inside organisation (e.g. local incident reports)

External event reporting systems:

• Reporting outside organisation (e.g. sharing with peers)



Mandatory event reporting systems

Mandatory reporting (to authorities) should ...

- ... focus on serious errors resulting in injury or death
 - ... ensure providers of medical care are held accountable to the public
 - ... require reporting of information in a standardised format to a national database





Mandatory event reporting systems

Two main purposes:

- ... to provide public with certain level of protection by assuring that most-serious errors are reported and investigated, and action is taken
 - to provide an incentive to hospitals to improve and invest in patient safety, helping to assure that hospitals offer comparable care





Mandatory event reporting systems

Filing of a report should not trigger a release of information:

- ... reporting should trigger an investigation
 - release of information should occur only after incident has been investigated thoroughly, and information released should be accurate and verified
 - employees should feel confident that response to reporting of significant error will be reasonable and justified





Mandatory event reporting systems

Radiotherapy: A mix of radiation and medicine

- Legislation and regulations concerning reporting of incidents in radiotherapy can be covered in relation to radiation protection and/or health
- In some countries, radiation protection regulations make it mandatory to report radiotherapy incidents to a regulatory authority
- In some countries, health regulations make it mandatory to report radiotherapy incidents to another regulatory authority
- Some countries stipulate that local recording of incidents is mandatory.
 Potential incidents are covered in some countries





Voluntary event reporting

Voluntary event reporting systems

Voluntary reporting should ...

- ... focus on errors that result in little or no harm to patients
 - encourage hospitals to focus on improvement of safety environment
 - have mechanisms to ensure that information and lessons learned can be shared effectively





Voluntary event reporting

Voluntary event reporting systems

Voluntary reporting should ...

... have mechanisms that allow for anonymous reporting of errors or circumstances that could lead to errors, and allow handling in confidence

Staff reporting should not fear punishment





Internal event reporting

Internal event reporting systems

Reporting of incidents within organisation

- Specific in relation to intra-organisation ...
 - ... procedures
 - ... equipment
 - … characteristics
 - "Lessons to learn" become more direct and explicit
- Follows up management of actual patients affected by the incidents



Category	Description	Sub-Description	Frequency
Incident information	Description of incident		25 of 27
	Cause of incident		9 of 27
	Number of fractions affected		10 of 27
	When did it occur?	Date	18 of 27
		Time	12 of 27
		Weekday	1 of 27
	Detection of incident	How	4 of 27
		By whom	2 of 27
		Where in process	1 of 27
		Date	3 of 27
	Estimate of deviation	Absorbed dose	2 of 27
		Dose after correction	2 of 27
		Field location	1 of 27
		Correctable or not	3 of 27
	Clinical significance or risk to patient		12 of 27
	Contributing factors	General comment	4 of 27
		Treatment plan complexity	1 of 27
		Staffing levels	4 of 27
		Staffing composition	2 of 27
		Staff on leave	1 of 27
		Distractions	1 of 27



Category	Description	Sub-Description	Frequency
Action information	Corrective action	Action to be performed and/or already taken	22 of 27
		Responsible for this	3 of 27
		Date for completion	5 of 27
	Preventive action	Recommended action to prevent recurrence	10 of 27
		Procedural changes	2 of 27
		Confirmation of preventive action	3 of 27
	Communication	Patient informed	4 of 27
		Responsible physician informed	13 of 27
		Authority informed	9 of 27
		General	6 of 27

Source: Radiation Oncology Safety Information System, unpublished survey, 2002.



External event reporting

External event reporting systems

Reporting of incidents outside organisation

- "Lessons to learn" will come from a bigger pool of events
- An incident in another hospital can lead to identification of the hazard before a similar incident is realised in your own hospital
- More extensive pool of events → better identification of safetycritical steps in the radiotherapy process where errors are likely to occur or be detected
 - A general culture of safety awareness can be created by making information available on details of incidents, near-incidents and corrective actions



Inadvertent loss of wedge code information (ROSIS report #284)

Due to the breakdown of a linear accelerator, a patient was moved to another accelerator for a single fraction.

As an inherent part of the design of the R&V system, the wedge information in the R&V system was not transferred automatically to the new treatment unit.

The wedge code was manually input properly for the single fraction at the second unit, but when the patient was transferred back to the original unit, the wedge code was not put in again.

As a result, the patient received treatment without wedges for three fractions before discovery, causing accidental delivery of the incorrect absorbed dose and dose distribution.



Inadvertent loss of wedge code information (ROSIS report #284)

Due to the breakdown of a linear accelerator, a patient was moved to another accelerator for a single fraction.

As an inherent part of the design of the R&V system, the wedge information in the R&V system was not trafficiented automatically to the new treatment unit.

The wedge code was manually input properly for the single fraction at the second unit, but when the patient was transferred back to the original unit, the wedge code was not put in again.

As a result, the patient received treatment without wedges for three fractions before discovery, causing accidental delivery of the incorrect absorbed dose and dose distribution.



Inadvertent rotational treatment of a patient (ROSIS report #284)

During the first treatment of a patient with an electron field, it was noted that the gantry started to rotate.

The prescription was for static treatment, not rotational. An error had been made when preparing the R&V entry of the treatment, where a checkbox had been accidentally checked for rotational treatment.

It was also noted in another report to ROSIS (Incident Report #689) that, for this particular type of R&V system, the checkbox for rotational treatment on the screen was placed near the icon for closing the window after finalising the R&V entry, leading to inadvertent activation of rotational treatment.



Inadvertent rotational treatment of a patient (ROSIS report #284)

During the first treatment of a patient with an electron field, it was noted that the gantry started to rotate.

The prescription was for static treatment, not rotational. An error had been made when preparing the B&V entrarene Saturent, where a checkbox had been accident acking for rotations treatment.

It was also noted in anotand all to ROSIS (Incident Report #589) that, for this particular type of R&V system, the check officient all treatment on the screen was placed near the icon for the design finalising the R&V entry, leading to inadverse termination of rotational treatment.



Common terminology for event reporting systems would be of value

Severity classification; Causes / contributing factors classification; Standardized process map; Other terminology



Incidents and near-misses

Accident:

Any unintended event, including operating errors, equipment failures and other mishaps, the consequences or potential consequences of which are not negligible from the point of view of protection or safety.

Incident:

Any unintended event, including operating errors, equipment failures, initiating events, accident precursors, near misses or other mishaps, or unauthorized act, malicious or non-malicious, the consequences or potential consequences of which are not negligible from the point of view of protection or safety.

(Source: IAEA Safety Glossary, 2007)




Incidents are important

Variable magnitude:

Many incidents (e.g. mistake in calculation of monitor units for a single patient) can have a variable magnitude (e.g. for Patient 1, the mistake causes a dose deviation of 5%, while for Patient 2, the same type of mistake causes a dose deviation of 50%).





Incidents are important

More events:

Incidents are more numerous than accidents, so there are more opportunities to learn and improve the safety, than by only looking at major accidents.





consequence ranges		
catastrophic		
critical		
marginal		
negligible		



likelihood ranges / consequence ranges	improbable 10-6	remote 10-5	occasional 10-3	probable 10-2
catastrophic				
critical				
marginal				
negligible				



<u>likelihood ranges</u> / consequence ranges	improbable 10-6	remote 10-5	occasional 10-3	probable 10-2
catastrophic	III	I	I	I.
critical	IV	III		L I
marginal	IV	IV		=
negligible	IV	IV	IV	III



<u>likelihood ranges</u> / consequence ranges	improbable 10-6	remote 10-5	occasional 10-3	probable 10-2
catastrophic	III	l		I
critical	IV	III		I
marginal	IV	IV		I
negligible	IV	IV	IV	III

risk rank	CATEGORY
I.	unacceptable
I	undesirable
ш	acceptable with controls
IV	acceptable as is



• Risk ranking matrix:

2

<u>likelihood ranges</u> / consequence ranges	imp	orebable	10 ⁻⁶	remote 10⁵	occasional 10-3	probable 10-2
catastrophic		III		II	I	I
critical					I	I
marginal		IV		IV	III	II
negligible		IV		IV	IV	III

Major systematic	<u>risk rank</u>	CATEGORY
accident	I	unacceptable
	II	undesirable
	III	acceptable with controls
	IV	acceptable as is
IACA		

<u>likelihood ranges</u> / <u>consequence ranges</u>	improbable 10-6	remote 10-5	occasional 10-3	probable 10-2
catastrophic		I	I	I
critical	IV	III	-	I
marginal	IV	IV		
negligible	IV	ÏV	IV	→ (III)

Non-systematic	<u>risk rank</u>	CATEGORY
incident	I	unacceptable
	II	undesirable
		acceptable with controls
	IV	acceptable as is

 Independent calculation checks monitored between 1998 and 2003 (27830 charts / treatment plans were checked)



 In total, 4.3% of charts / treatment plans had mistakes found at some point: either prior to treatment or when treatment had started



 Independent calculation checks monitored between 1998 and 2003 (27830 charts / treatment plans were checked)



 The first check found mistakes in 3.5% of all charts / treatment plans – 0.8% remained



 Independent calculation checks monitored between 1998 and 2003 (27830 charts / treatment plans were checked)



 The second check found mistakes in 0.5% of all charts / treatment plans – 0.3% remained



 Independent calculation checks monitored between 1998 and 2003 (27830 charts / treatment plans were checked)



 The second check found mistakes in 0.5% of all charts / treatment plans – 0.3% remained



 Independent calculation checks monitored between 1998 and 2003 (27830 charts / treatment plans were checked)



Safety Improvement Initiatives

A good city wall with properly built bastions can be effective







