

**ICTP School On Medical Physics For
Radiation Therapy:
Dosimetry And Treatment Planning For Basic And Advanced
Applications**
13 - 24 April 2015
Miramare, Trieste, Italy

Dosimetry Exercise

G. Hartmann
EFOMP & German Cancer Research Center (DKFZ)
g.hartmann@dkfz.de

Exercise: Beam calibration at a 15 MV photon beam

Remark 1: Beam calibration here means:

**Determination of absorbed dose to water per 100
monitor units in a water phantom at reference
conditions**

**using
the IAEA Code of Practice TRS398**

Remark 2: Use Excel for calculation and plotting

Introduction: Formalism

- The absorbed dose to water at the reference depth z_{ref} in water for a reference beam of quality Q_0 and in the absence of the chamber is given by:

$$D_{w,Q} = M_Q N_{D,w,Q_0} k_{Q,Q_0}$$

Discussion of the meaning of the three quantities

$$D_{w,Q} = M_Q N_{D,w,Q_0} k_{Q,Q_0}$$

 M_Q

is the **chamber reading in beam of quality Q** and corrected for influence quantities to the reference conditions used in the standards laboratory.

 N_{D,w,Q_0}

is the **calibration factor** in terms of absorbed dose to water of the dosimeter obtained from a standards laboratory

 k_{Q,Q_0}

is the **beam quality factor** which corrects for the effects of the difference between the reference beam quality Q_0 and the actual user quality Q


We are using a virtual equipment

- ❑ High Energy Photon Beam:
15 MV accelerating potential

- ❑ Ionization chamber:
PTW Farmer Type 30013



- ❑ water phantom:




Virtual Equipment

further equipment:

- electrometer
- thermometer
- barometer

Virtual Equipment: Simulation Program




Objectives:

1. To learn of how to set up the measuring equipment
2. To be able to differentiate between a depth dose measurement and a calibration measurement
3. To know how a charge measurement obtained by using some monitor units has to be manually converted into dose in water per 100 MU under reference condition

**Some more details on the ionization chamber type to
be used for beam calibration:
PTW Farmer Type 30013**

Calibration factor:	$N = 5.233 \text{ Gy/C}$
Radius of sensitive volume:	$r = 3.1 \text{ mm}$
Voltage to be applied:	400 V
Polarity:	as used with calibration

Main steps to be performed:

1. Prepare the virtual accelerator:
 - set gantry angle at zero
 - set collimator angle at zero

select angle and
press start continuously

 - select type of radiation and energy
 - select reference field size
2. Prepare water phantom:
 - needs water filling
 - needs adjustment of water surface to laser lines
(laser lines may be switched on with OPTIONS)
 - measure temperature and air pressure
(see Environment, utmost right

Main steps to be performed (cont):

3. Prepare chamber:
 - adjust reference point of chamber to central ray
 - position the chamber correctly to zero depth
 - voltage and polarity
4. Determine the quality index Q
 - determine a PDD and use the depth dose method
5. Determine the quality correction factor
 - use interpolation between table values
6. Determine charge under reference conditions
 - measure charge
 - apply correction factors
7. Finally obtain calibration value

Note:

In high energy beams, cylindrical chambers are used for both, for

- a) depth dose measurements
- b) calibration measurements

Thus depth dose measurements and beam calibration can be performed with a single chamber type.

However, they must be positioned in different ways:

- a) for depth dose: **effective point** at measuring depth
- b) for calibration: **central axis** at measuring depth

Note:

Depth dose measurements with this virtual accelerator are performed in the following way:

Purpose of Measurement:
☐ Calibration ☒ Depth Dose Measurement

Depth Dose Measurement Parameters

Start depth: cm
Stop depth: cm
Step increment: cm

Monitor Preselection: MU

Measure

The screenshot shows a software interface for depth dose measurements. It has a section for 'Purpose of Measurement' with two radio buttons: 'Calibration' and 'Depth Dose Measurement' (which is selected). Below this is a section titled 'Depth Dose Measurement Parameters' containing three input fields: 'Start depth', 'Stop depth', and 'Step increment', each followed by a unit of 'cm'. At the bottom of this section is a 'Monitor Preselection' field with a unit of 'MU'. A 'Measure' button is located at the bottom right of the form. Four arrows point from yellow callout boxes to the 'Start depth' field, the 'Stop depth' field, the 'MU' field, and the 'Measure' button.

Start depth must be greater than 0.5

Stop depth must be greater than start depth


MU required for each single depth (no continuous measurement)

Results can be copied and paste into an EXCEL fle


Example of a depth dose measurement at central ray

Water Temperature: 20.60 °C

Air Pressure: 98.18 kPa



Step 4: Determination of the quality index for HE photons the depth dose method:



$$\Rightarrow M_{10} = 7.745 \text{ nC}$$

$$\Rightarrow M_{20} = 5.002 \text{ nC}$$

$$PDD_{20,10} = \frac{M_{20}}{M_{10}} = 0.652$$

Determination of the quality index for HE photons:

$$PDD_{20,10} = 0.65$$

$$TPR_{20,10} = 1.2661 \cdot PDD_{20,10} - 0.0595$$

$$TPR_{20,10} = 0.767$$

Step 5: Determination of k_Q

□ reference beam quality

Co-60

used beam quality?? (15 MV)

Requires input of the beam quality index Q

Step 5: Determination of k_Q

IAEA TRS 398
CALCULATED VALUES OF k_Q FOR HIGH-ENERGY PHOTON
BEAMS, FOR VARIOUS CYLINDRICAL IONIZATION CHAMBERS AS
A FUNCTION OF BEAM QUALITY $TPR_{20,10}$

Quality index	0.74	0.76	0.767	0.78	0.80
PTW 30006/30013	0.98	0.975		0.968	0.960

by linear interpolation: 0.973

Charge measurement under
reference conditions
for 15 MV photons

- ❑ field size: 10 cm x 10 cm
- ❑ SSD: 100 cm
- ❑ phantom: water phantom
- ❑ measurement depth
in water: 10 cm
- ❑ positioning of
chamber: central electrode at
measuring depth

Example of Measurement at a depth of 10 cm:

Measure

Your Measured Results:

SSD =99.98 cm	d = 10.00 cm	M = 7.670 nC
SSD =99.98 cm	d = 10.00 cm	M = 7.644 nC
SSD =99.98 cm	d = 10.00 cm	M = 7.668 nC
SSD =99.98 cm	d = 10.00 cm	M = 7.669 nC
SSD =99.98 cm	d = 10.00 cm	M = 7.674 nC

Monitor Preselection: **MU**

Measure

depth in cm

10.0

Mean charge in nC

7.674

Calibration Certificate

Reference conditions : T_0 : 20.0 °C p_0 : 101.325 kPa R.H.: 50 %

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, which for a normal distribution provides a level of confidence of approximately 95%.

The secondary standard of this laboratory is traceable to the PTB in Braunschweig (German Federal Institute of Physics and Metrology).

Calibration reported in this certificate was carried out in accordance with the procedures described in the IAEA TRS 398 Code of Practice.

Measuring conditions: Phantom size : 30 cm × 30 cm × 30 cm
 Phantom material : water
 Source to phantom surface distance (SSD) : 100 cm
 Field size at the phantom surface : 10 cm × 10 cm
 Depth in phantom of the reference point of the chamber : 5 g·cm⁻²
 Reference point of the IC : on the chamber axis at the centre of the cavity volume
 Chamber orientation : the beam axis perpendicular to the chamber axis
 If the chamber stem has a mark, the mark is oriented towards the radiation source
 Waterproof sleeve (PMMA) : NO
 Sleeve Serial Number: -
 Polarizing potential of collecting (central) electrode : 400 V
 Dose rate : 1.0 Gy·min⁻¹

Recombination correction has not been applied

Date of calibration

Head of the Dosimetry Laboratory

Calibration performed by

28.04.2006

[Handwritten signature]

Measurement under **reference conditions**

- reference water temperature $T_0 = 20^\circ\text{C}$
- reference air pressure (absolute!!!) $P_0 = 101.325 \text{ kPa}$

measured water temperature: $T = 20.6^\circ\text{C}$

measured air pressure (absolute!!!): $P = 98.18 \text{ kPa}$

air density correction:

multiply measured M with:

$$\frac{(273.2 + T)}{(273.2 + T_0)} \frac{P_0}{P} = 1.034$$

Measurement under **reference conditions**

□ reference saturation 100%

used polarizing potential: 400 V
saturation is 100% ???

measure charge under
identical conditions with
the lower voltage of 100 V

voltage	charge in nC
400.0	7.674
100.0	7.587

Measurement under **reference conditions**

□ reference saturation 100%

$$k_s = a_0 + a_1 \left(\frac{M_1}{M_2} \right) + a_2 \left(\frac{M_1}{M_2} \right)^2$$

TABLE 4.VII. QUADRATIC FIT COEFFICIENTS, FOR THE CALCULATION OF k_s BY THE “TWO-VOLTAGE” TECHNIQUE IN PULSED AND PULSED-SCANNED RADIATION, AS A FUNCTION OF THE VOLTAGE RATIO V_1/V_2 [76]

V_1/V_2	Pulsed			Pulsed scanned		
	a_0	a_1	a_2	a_0	a_1	a_2
2.0	2.337	-3.636	2.299	4.711	-8.242	4.533
2.5	1.474	-1.587	1.114	2.719	-3.977	2.261
3.0	1.198	-0.875	0.677	2.001	-2.402	1.404
3.5	1.080	-0.542	0.463	1.665	-1.647	0.984
4.0	1.022	-0.363	0.341	1.468	-1.200	0.734
5.0	0.975	-0.188	0.214	1.279	-0.750	0.474

Measurement under **reference conditions**

□ reference saturation 100%

correction for the ion recombination effect:

multiply measured M with:

$$a_o + a_1 \left(\frac{M_1}{M_2} \right) + a_2 \left(\frac{M_1}{M_2} \right)^2 = 1.004$$

Measurement under **reference conditions**

□ reference polarity

????

used polarizing potential:
polarity effect ???

+400 V

Measurement uncertainty :

$U = 2.2 \%$

Reference conditions :

$T_0 : 20.0 \text{ }^\circ\text{C}$

$p_0: 101.325 \text{ kPa}$

R.H.: 50 %

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, which for a normal distribution provides a level of confidence of approximately 95%.

The secondary standard of this laboratory is traceable to the PTB in Braunschweig (German Federal Institute of Physics and Metrology).

Calibration reported in this certificate was carried out in accordance with the procedures described in the IAEA TRS 398 Code of Practice.

Measuring conditions:

Phantom size : 30 cm × 30 cm × 30 cm

Phantom material : water

Source to phantom surface distance (SSD) : 100 cm

Field size at the phantom surface : 10 cm × 10 cm

Depth in phantom of the reference point of the chamber : 5 g·cm⁻²

Reference point of the IC : on the chamber axis at the centre of the cavity volume

Chamber orientation : the beam axis perpendicular to the chamber axis

If the chamber stem has a mark, the mark is oriented towards the radiation source

Waterproof sleeve (PMMA) : NO

Sleeve Serial Number: -

Polarizing potential of collecting (central) electrode : 400 V

Dose rate : 1.0 Gy·min⁻¹

Recombination correction has not been applied

Date of calibration

Head of the Dosimetry Laboratory

Calibration performed by

28.04.2006

Measurement under **reference conditions**

□ reference polarity ????

used polarizing potential: +400 V
polarity effect ???

The polarity effect for photon beams usually is very small.

In such a case where no information on the polarity used at calibration is given, it is better **not** to perform any correction. It may be a wrong correction!

Step 6: Determination of corrected charge M_Q

chamber reading in beam of quality Q

corrected for influence quantities to the reference conditions

$$M_Q = 7.674 \cdot 1.034 \cdot 1.004 = 7.967 \text{ [nC/50 MU]}$$

k_{TP}

k_s

Determination of

$$N_{D,w,Q_0}$$

Calibration Certificate	
Calibration laboratory for ionising radiation quantities	Calibration mark
	000877
	04-06
Object :	
Manufacturer :	
Type :	
Serial number :	
Beam quality :	Co-60
Absorbed dose to water calibration factor :	$N_{D,w} = 5.233 \times 10^7 \text{ Gy/C}$

Using the formalism

$$D_{w,Q} = M_Q N_{D,w,Q_0} k_{Q,Q_0}$$

$$M_Q = 7.967 \text{ nC/50 MU}$$

$$N_{D,w,Q_0} = 5.233 \times 10^7 \text{ Gy/C}$$

$$k_{Q,Q_0} = 0.973$$

$$D_{w,15\text{MV}} = 0.812 \text{ Gy/100 MU}$$